

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00852	Calib. Date:	09/12/2018
Certification File No.:	00853	Cert. No.:	00033
Linearity File No.:	00854	Cert. Date:	09/12/2018
Solution File No.:	00858	Lin. No.:	00027
Sequential File No.:	00858	Lin. Date:	09/12/2018
		Soln. Date:	12/19/2018
		Soln. No.:	00222
		File Date:	12/19/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0135
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0388

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:08S		
Control 1 EC	0.101%	01:09S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:09S		
Control 2 EC	0.099%	01:10S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:11S		
Control 3 EC	0.099%	01:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:12S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

Badge No.: 321

Date: 12/19/2018

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
Location: LONGPORT POLICE  
Calibration File No.: 00852 Calib. Date: 09/12/2018 Calib. No.: 00033  
Certification File No.: 00853 Cert. Date: 09/12/2018 Cert. No.: 00027  
Linearity File No.: 00854 Lin. Date: 09/12/2018 Lin. No.: 00027  
Solution File No.: 00857 Soln. Date: 11/13/2018 Soln. No.: 00221  
Sequential File No.: 00857 File Date: 11/13/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135  
Control Solution %: 0.100% Expires: 01/31/2020  
Solution Control Lot: 18050 Bottle No.: 0157

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:25S		
Control 1 EC	0.100%	11:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26S		
Control 2 EC	0.099%	11:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28S		
Control 3 EC	0.099%	11:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: David Cincotta 321

Badge No.: 321

Date: 11/13/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00852	Calib. Date: 09/12/2018
Certification File No.:	00853	Calib. No.: 00033
Linearity File No.:	00854	Cert. Date: 09/12/2018
Solution File No.:	00856	Cert. No.: 00027
Sequential File No.:	00856	Lin. Date: 09/12/2018
		Lin. No.: 00027
		Soln. Date: 10/13/2018
		Soln. No.: 00220
		File Date: 10/13/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0161

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:24D		
Control 1 EC	0.098%	02:25D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:25D		
Control 2 EC	0.097%	02:26D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:27D		
Control 3 EC	0.097%	02:27D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature:  321

Badge No.: 321

Date: 10/13/2018

# Calibrating Unit New Standard Solution Report

COPY

**Equipment** Alcotest 7110 MKIII-C  
**Location:** LONGPORT POLICE  
**Calibration File No.:** 00852  
**Certification File No.:** 00853  
**Linearity File No.:** 00854  
**Solution File No.:** 00855  
**Sequential File No.:** 00855

Serial No.: ARXA-0067

**Calib. Date:** 09/12/2018      **Calib. No.:** 00033  
**Cert. Date:** 09/12/2018      **Cert. No.:** 00027  
**Lin. Date:** 09/12/2018      **Lin. No.:** 00027  
**Soln. Date:** 09/12/2018      **Soln. No.:** 00219  
**File Date:** 09/12/2018

**Calibrating Unit:** WET  
**Control Solution %:** 0.100%  
**Solution Control Lot:** 18050

Model No.: CU-34

Serial No.: DDXC S3-0135  
**Expires:** 01/31/2020  
**Bottle No.:** 0092

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:58D		
Control 1 EC	0.101%	14:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:59D		
Control 2 EC	0.100%	15:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00D		
Control 3 EC	0.099%	15:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: \_\_\_\_\_

DDWA P2-199 (KA)

**Changed By:**

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: \_\_\_\_\_

Sgt. K. W. #6704

Badge No.: 6704

Date: 09/12/2018

original with  
Calibration Record

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00837	Calib. Date: 03/28/2018
Certification File No.:	00838	Calib. No.: 00032
Linearity File No.:	00839	Cert. Date: 03/28/2018
Solution File No.:	00850	Cert. No.: 00026
Sequential File No.:	00850	Lin. Date: 03/28/2018
		Lin. No.: 00026
		Soln. Date: 08/17/2018
		Soln. No.: 00218
		File Date: 08/17/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0321

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:58D		
Control 1 EC	0.098%	12:59D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	12:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59D		
Control 2 EC	0.097%	13:00D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	13:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01D		
Control 3 EC	0.096%	13:01D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	13:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

*David Cincotta 321*

Badge No.: 321

Date: 08/17/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00837	Calib. Date: 03/28/2018
Certification File No.:	00838	Calib. No.: 00032
Linearity File No.:	00839	Cert. Date: 03/28/2018
Solution File No.:	00849	Cert. No.: 00026
Sequential File No.:	00849	Lin. Date: 03/28/2018
		Lin. No.: 00026
		Soln. Date: 08/17/2018
		Soln. No.: 00217
		File Date: 08/17/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0104

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:13D		
Control 1 EC	0.095%	11:13D	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.094%	11:13D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	11:14D		
Control 2 EC	0.095%	11:15D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	11:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15D		
Control 3 EC	0.095%	11:16D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.095%	11:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

*David Cincotta* 321

Badge No.: 321

Date: 08/17/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00837	Calib. Date: 03/28/2018
Certification File No.:	00838	Calib. No.: 00032
Linearity File No.:	00839	Cert. Date: 03/28/2018
Solution File No.:	00848	Cert. No.: 00026
Sequential File No.:	00848	Lin. Date: 03/28/2018
		Lin. No.: 00026
		Soln. Date: 07/17/2018
		Soln. No.: 00216
		File Date: 07/17/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0108

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:36D		
Control 1 EC	0.097%	22:37D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	22:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:38D		
Control 2 EC	0.096%	22:38D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	22:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:39D		
Control 3 EC	0.096%	22:40D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.095%	22:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SILVA

SGT. 

First Name: JAMES

MI: F

Signature: \_\_\_\_\_

Badge No.: 319

Date: 07/17/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00837	Calib. Date: 03/28/2018
Certification File No.:	00838	Calib. No.: 00032
Linearity File No.:	00839	Cert. Date: 03/28/2018
Solution File No.:	00846	Cert. No.: 00026
Sequential File No.:	00846	Lin. Date: 03/28/2018
		Lin. No.: 00026
		Soln. Date: 07/07/2018
		Soln. No.: 00215
		File Date: 07/07/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0330

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:37D		
Control 1 EC	0.096%	20:37D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.095%	20:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:38D		
Control 2 EC	0.095%	20:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	20:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:39D		
Control 3 EC	0.095%	20:40D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	20:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

*David Cincotta* 321

Badge No.: 321

Date: 07/07/2018



# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
Location: LONGPORT POLICE  
Calibration File No.: 00837 Calib. Date: 03/28/2018 Calib. No.: 00032  
Certification File No.: 00838 Cert. Date: 03/28/2018 Cert. No.: 00026  
Linearity File No.: 00839 Lin. Date: 03/28/2018 Lin. No.: 00026  
Solution File No.: 00845 Soln. Date: 06/14/2018 Soln. No.: 00214  
Sequential File No.: 00845 File Date: 06/14/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0154  
Control Solution %: 0.100% Expires: 01/31/2020  
Solution Control Lot: 18050 Bottle No.: 0275

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:51D		
Control 1 EC	0.096%	06:52D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	06:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:53D		
Control 2 EC	0.095%	06:53D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	06:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:54D		
Control 3 EC	0.095%	06:55D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	06:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SILVA

First Name: JAMES

MI: F

Signature: \_\_\_\_\_



Badge No.: 319

Date: 06/14/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00837	Calib. Date: 03/28/2018
Certification File No.:	00838	Cert. No.: 00032
Linearity File No.:	00839	Cert. Date: 03/28/2018
Solution File No.:	00842	Lin. No.: 00026
Sequential File No.:	00842	Soln. Date: 05/24/2018
		Soln. No.: 00213
		File Date: 05/24/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0271

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:04D		
Control 1 EC	0.096%	14:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	14:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:05D		
Control 2 EC	0.095%	14:06D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	14:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:07D		
Control 3 EC	0.095%	14:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	14:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature:  # 321

Badge No.: 321

Date: 05/24/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00837	Calib. Date: 03/28/2018
Certification File No.:	00838	Calib. No.: 00032
Linearity File No.:	00839	Cert. Date: 03/28/2018
Solution File No.:	00841	Cert. No.: 00026
Sequential File No.:	00841	Lin. Date: 03/28/2018
		Lin. No.: 00026
		Soln. Date: 04/26/2018
		Soln. No.: 00212
		File Date: 04/26/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0389

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:11D		
Control 1 EC	0.097%	04:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:13D		
Control 2 EC	0.096%	04:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	04:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:14D		
Control 3 EC	0.096%	04:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	04:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

*David Cincotta* 321

Badge No.: 321

Date: 04/26/2018

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
Location: LONGPORT POLICE  
Calibration File No.: 00827 Calib. Date: 10/20/2017 Calib. No.: 00031  
Certification File No.: 00828 Cert. Date: 10/20/2017 Cert. No.: 00025  
Linearity File No.: 00829 Lin. Date: 10/20/2017 Lin. No.: 00025  
Solution File No.: 00836 Soln. Date: 03/14/2018 Soln. No.: 00210  
Sequential File No.: 00836 File Date: 03/14/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135  
Control Solution %: 0.100% Expires: 01/31/2020  
Solution Control Lot: 18050 Bottle No.: 0105

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:27S		
Control 1 EC	0.097%	09:28S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:28S		
Control 2 EC	0.098%	09:29S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30S		
Control 3 EC	0.098%	09:30S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: David Cincotta 321

Badge No.: 321

Date: 03/14/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00827	Calib. Date:	10/20/2017
Certification File No.:	00828	Cert. No.:	00031
Linearity File No.:	00829	Cert. Date:	10/20/2017
Solution File No.:	00835	Lin. No.:	00025
Sequential File No.:	00835	Soln. Date:	02/10/2018
		Soln. No.:	00209
		File Date:	02/10/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0135
Solution Control Lot:	17050	Expires:	01/30/2019
		Bottle No.:	1247

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:04S		
Control 1 EC	0.097%	09:04S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	09:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:05S		
Control 2 EC	0.096%	09:06S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	09:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:07S		
Control 3 EC	0.095%	09:07S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	09:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: \_\_\_\_\_

Badge No.: 321

Date: 02/10/2018

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
Location: LONGPORT POLICE  
Calibration File No.: 00827 Calib. Date: 10/20/2017 Calib. No.: 00031  
Certification File No.: 00828 Cert. Date: 10/20/2017 Cert. No.: 00025  
Linearity File No.: 00829 Lin. Date: 10/20/2017 Lin. No.: 00025  
Solution File No.: 00833 Soln. Date: 01/19/2018 Soln. No.: 00208  
Sequential File No.: 00833 File Date: 01/19/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135  
Control Solution %: 0.100% Expires: 01/30/2019  
Solution Control Lot: 17050 Bottle No.: 1384

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:48S		
Control 1 EC	0.102%	02:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:50S		
Control 2 EC	0.100%	02:50S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:51S		
Control 3 EC	0.099%	02:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature: David Cincotta 321

Badge No.: 321

Date: 01/19/2018