

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
 Location: LONGPORT POLICE  
 Calibration File No.: 00876 Calib. Date: 07/22/2019 Calib. No.: 00035  
 Certification File No.: 00877 Cert. Date: 07/22/2019 Cert. No.: 00029  
 Linearity File No.: 00878 Lin. Date: 07/22/2019 Lin. No.: 00029  
 Solution File No.: 00892 Soln. Date: 11/28/2019 Soln. No.: 00235  
 Sequential File No.: 00892 File Date: 11/28/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135  
 Control Solution %: 0.100% Expires: 01/31/2020  
 Solution Control Lot: 18050 Bottle No.: 0320


Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:01S		
Control 1 EC	0.097%	08:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:02S		
Control 2 EC	0.097%	08:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:04S		
Control 3 EC	0.098%	08:05S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA First Name: DAVID MI: P

Signature:  321  
 Badge No.: 321  
 Date: 11/28/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135  
 Control Solution %: 0.100% Expires: 01/31/2020  
 Solution Control Lot: 18050 Bottle No.: 0320

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:01S		
Control 1 EC	0.097%	08:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:02S		
Control 2 EC	0.097%	08:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:03S	34.0°C	*** TEST PASSED ***

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00876	Calib. Date:	07/22/2019
Certification File No.:	00877	Cert. Date:	07/22/2019
Linearity File No.:	00878	Lin. Date:	07/22/2019
Solution File No.:	00891	Soln. Date:	11/27/2019
Sequential File No.:	00891	File Date:	11/27/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0135
Control Solution %:	0.100%			Expires:	01/31/2020
Solution Control Lot:	I8050			Bottle No.:	0320

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:03S		
Control 1 EC	0.099%	23:03S	33.9°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	23:03S	33.9°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	0.015%	23:04S		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature:  321

Badge No.: 321

Date: 11/27/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0135
Control Solution %:	0.100%			Expires:	01/31/2020
Solution Control Lot:	I8050			Bottle No.:	0320

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:03S		
Control 1 EC	0.099%	23:03S	33.9°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	23:03S	33.9°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	0.015%	23:04S		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	

# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00876	Calib. Date:	07/22/2019
Certification File No.:	00877	Cert. Date:	07/22/2019
Linearity File No.:	00878	Lin. Date:	07/22/2019
Solution File No.:	00887	Soln. Date:	10/17/2019
Sequential File No.:	00887	File Date:	10/17/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0135
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0386

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:21D		
Control 1 EC	0.100%	07:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:22D		
Control 2 EC	0.099%	07:23D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:23D		
Control 3 EC	0.100%	07:24D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA	First Name: DAVID	MI: P
Signature: 	Badge No.: 321	Date: 10/17/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0135
Control Solution %:	0.100%			Expires:	01/31/2020
Solution Control Lot:	18050			Bottle No.:	0386

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:21D		
Control 1 EC	0.100%	07:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:22D		
Control 2 EC	0.099%	07:23D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:23D	33.9°C	*** TEST PASSED ***

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00876	Calib. Date:	07/22/2019
Certification File No.:	00877	Cert. Date:	07/22/2019
Linearity File No.:	00878	Lin. Date:	07/22/2019
Solution File No.:	00886	Soln. Date:	09/20/2019
Sequential File No.:	00886	File Date:	09/20/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0135
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0166

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:18D		
Control 1 EC	0.100%	09:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:19D		
Control 2 EC	0.099%	09:20D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:21D		
Control 3 EC	0.099%	09:21D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:22D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature:  321

Badge No.: 321

Date: 09/20/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0135
Control Solution %:	0.100%			Expires:	01/31/2020
Solution Control Lot:	18050			Bottle No.:	0166

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:18D		
Control 1 EC	0.100%	09:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:19D		
Control 2 EC	0.099%	09:20D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:20D	33.9°C	*** TEST PASSED ***

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00876	Calib. Date:	07/22/2019
Certification File No.:	00877	Cert. No.:	00035
Linearity File No.:	00878	Cert. Date:	07/22/2019
Solution File No.:	00881	Lin. Date:	07/22/2019
Sequential File No.:	00881	Lin. No.:	00029
		Soln. Date:	08/19/2019
		Soln. No.:	00231
		File Date:	08/19/2019


Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0135
Control Solution %:	0.100%			Expires:	01/31/2020
Solution Control Lot:	18050			Bottle No.:	0319

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:47D		
Control 1 EC	0.100%	11:47D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48D		
Control 2 EC	0.099%	11:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:50D		
Control 3 EC	0.099%	11:50D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA	First Name: DAVID	MI: P
Signature: 	Badge No.: 321	
	Date: 08/19/2019	

Sequential File No.: 00881      File Date: 08/19/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0135
Control Solution %:	0.100%			Expires:	01/31/2020
Solution Control Lot:	18050			Bottle No.:	0319

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:47D		
Control 1 EC	0.100%	11:47D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48D		
Control 2 EC	0.099%	11:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:49D	34.0°C	*** TEST PASSED ***

# Calibrating Unit New Standard Solution Report



**COPY**

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
Location: LONGPORT POLICE  
Calibration File No.: 00876 Calib. Date: 07/22/2019 Calib. No.: 00035  
Certification File No.: 00877 Cert. Date: 07/22/2019 Cert. No.: 00029  
Linearity File No.: 00878 Lin. Date: 07/22/2019 Lin. No.: 00029  
Solution File No.: 00879 Soln. Date: 07/22/2019 Soln. No.: 00230  
Sequential File No.: 00879 File Date: 07/22/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135  
Control Solution %: 0.100% Expires: 01/31/2020  
Solution Control Lot: 18050 Bottle No.: 0328

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:08D		
Control 1 EC	0.101%	10:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10D		
Control 2 EC	0.100%	10:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	10:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11D		
Control 3 EC	0.100%	10:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	10:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Temperature Probe Serial Number: DDWA P2-199 (DN)

### Changed By:

Last Name: NAPOLITANO

First Name: DAVID

MI: M.

Signature:  7237

Badge No.: 7237

Date: 07/22/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135  
Control Solution %: 0.100% Expires: 01/31/2020  
Solution Control Lot: 18050 Bottle No.: 0328

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:08D		
Control 1 EC	0.101%	10:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10D		
Control 2 EC	0.100%	10:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	10:11D	34.0°C	*** TEST PASSED ***

# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00861	Calib. Date: 01/31/2019
Certification File No.:	00862	Calib. No.: 00034
Linearity File No.:	00863	Cert. Date: 01/31/2019
Solution File No.:	00875	Cert. No.: 00028
Sequential File No.:	00875	Lin. Date: 01/31/2019
		Lin. No.: 00028
		Soln. Date: 07/21/2019
		Soln. No.: 00229
		File Date: 07/21/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0119

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:32D		
Control 1 EC	0.100%	03:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:33D		
Control 2 EC	0.098%	03:34D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:34D		
Control 3 EC	0.098%	03:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINGOTTA	First Name: DAVID	MI: P
Signature: <u> #321</u>	Badge No.: 321	Date: 07/21/2019

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXC S3-0154
Control Solution %:	0.100%		Expires: 01/31/2020
Solution Control Lot:	18050		Bottle No.: 0119

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:32D		
Control 1 EC	0.100%	03:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:33D		
Control 2 EC	0.098%	03:34D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:34D	34.0°C	*** TEST PASSED ***

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00861	Calib. Date:	01/31/2019
Certification File No.:	00862	Cert. Date:	01/31/2019
Linearity File No.:	00863	Lin. Date:	01/31/2019
Solution File No.:	00873	Soln. Date:	06/20/2019
Sequential File No.:	00873	File Date:	06/20/2019


Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0154
Control Solution %:	0.100%			Expires:	01/31/2020
Solution Control Lot:	18050			Bottle No.:	0326

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:41D		
Control 1 EC	0.101%	03:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:42D		
Control 2 EC	0.099%	03:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:43D		
Control 3 EC	0.100%	03:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA	First Name: DAVID	MI: P
Signature: <u> 321</u>	Badge No.: 321	Date: 06/20/2019

Sequential File No.:	00873	File Date:	06/20/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%		
Solution Control Lot:	18050	Serial No.:	DDXC S3-0154
		Expires:	01/31/2020
		Bottle No.:	0326

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:41D		
Control 1 EC	0.101%	03:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:42D		
Control 2 EC	0.099%	03:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:43D	34.0°C	*** TEST PASSED ***



# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00861	Calib. Date:	01/31/2019
Certification File No.:	00862	Cert. Date:	01/31/2019
Linearity File No.:	00863	Lin. Date:	01/31/2019
Solution File No.:	00872	Soln. Date:	05/18/2019
Sequential File No.:	00872	File Date:	05/18/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0154
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0272

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:47D		
Control 1 EC	0.099%	23:48D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:48D		
Control 2 EC	0.098%	23:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	23:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:50D		
Control 3 EC	0.097%	23:50D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA	First Name: DAVID	MI: P
Signature: 	Badge No.: 321	Date: 05/18/2019

Sequential File No.:	00872	File Date:	05/18/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0154
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0272

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:47D		
Control 1 EC	0.099%	23:48D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:48D		
Control 2 EC	0.098%	23:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	23:49D	33.9°C	*** TEST PASSED ***

# Calibrating Unit New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE	Calib. Date:	01/31/2019
Calibration File No.:	00861	Calib. No.:	00034
Certification File No.:	00862	Cert. Date:	01/31/2019
Linearity File No.:	00863	Lin. Date:	01/31/2019
Solution File No.:	00871	Soln. Date:	04/13/2019
Sequential File No.:	00871	File Date:	04/13/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0154
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0406

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:28D		
Control 1 EC	0.097%	21:28D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	21:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:29D		
Control 2 EC	0.097%	21:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:31D		
Control 3 EC	0.097%	21:31D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA	First Name: DAVID	MI: P
Signature: 	Badge No.: 321	Date: 04/13/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0154
Control Solution %:	0.100%	Expires:	01/31/2020	Bottle No.:	0406
Solution Control Lot:	18050				

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:28D		
Control 1 EC	0.097%	21:28D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	21:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:29D		
Control 2 EC	0.097%	21:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:30D	34.0°C	*** TEST PASSED ***

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00861	Calib. Date: 01/31/2019
Certification File No.:	00862	Calib. No.: 00034
Linearity File No.:	00863	Cert. Date: 01/31/2019
Solution File No.:	00867	Cert. No.: 00028
Sequential File No.:	00867	Lin. Date: 01/31/2019
		Lin. No.: 00028
		Soln. Date: 03/05/2019
		Soln. No.: 00225
		File Date: 03/05/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0411

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:00S		
Control 1 EC	0.099%	14:01S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02S		
Control 2 EC	0.100%	14:02S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:02S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:03S		
Control 3 EC	0.100%	14:04S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature:  321

Badge No.: 321

Date: 03/05/2019

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0411

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:00S		
Control 1 EC	0.099%	14:01S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02S		
Control 2 EC	0.100%	14:02S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:02S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:03S		

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE	Calib. Date:	09/12/2018
Calibration File No.:	00852	Calib. No.:	00033
Certification File No.:	00853	Cert. Date:	09/12/2018
Linearity File No.:	00854	Lin. Date:	09/12/2018
Solution File No.:	00860	Soln. Date:	01/24/2019
Sequential File No.:	00860	File Date:	01/24/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0135
Control Solution %:	0.100%			Expires:	01/31/2020
Solution Control Lot:	18050			Bottle No.:	0152

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:31S		
Control 1 EC	0.100%	19:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	19:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:32S		
Control 2 EC	0.097%	19:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:34S		
Control 3 EC	0.098%	19:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	19:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:35S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CINCOTTA

First Name: DAVID

MI: P

Signature:  321

Badge No.: 321

Date: 01/24/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0135
Control Solution %:	0.100%			Expires:	01/31/2020
Solution Control Lot:	18050			Bottle No.:	0152

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:31S		
Control 1 EC	0.100%	19:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	19:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:32S		
Control 2 EC	0.097%	19:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:33S	34.0°C	*** TEST PASSED ***