

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00905	Calib. Date: 07/02/2020
Certification File No.:	00906	Calib. No.: 00037
Linearity File No.:	00907	Cert. Date: 07/02/2020
Solution File No.:	00918	Cert. No.: 00031
Sequential File No.:	00918	Lin. Date: 07/02/2020
		Lin. No.: 00031
		Soln. Date: 12/02/2020
		Soln. No.: 00248
		File Date: 12/02/2020

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXC S3-0135
Control Solution %: 0.100%		Expires: 09/02/2022
Solution Control Lot: 20390		Bottle No.: 0257

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:26S		
Control 1 EC	0.099%	01:27S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:28S		
Control 2 EC	0.098%	01:29S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:29S		
Control 3 EC	0.099%	01:30S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 12/02/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00905	Calib. Date: 07/02/2020
Certification File No.:	00906	Calib. No.: 00037
Linearity File No.:	00907	Cert. Date: 07/02/2020
Solution File No.:	00917	Cert. No.: 00031
Sequential File No.:	00917	Lin. Date: 07/02/2020
		Lin. No.: 00031
		Soln. Date: 10/26/2020
		Soln. No.: 00247
		File Date: 10/26/2020

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	19300	Expires: 10/30/2021
		Bottle No.: 1237

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:14S		
Control 1 EC	0.100%	08:14S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:15S		
Control 2 EC	0.098%	08:16S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16S		
Control 3 EC	0.099%	08:17S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SILVA

First Name: JAMES

MI: F

Signature: _____



Badge No.: 319

Date: 10/26/2020

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00905	Calib. Date: 07/02/2020
Certification File No.:	00906	Calib. No.: 00037
Linearity File No.:	00907	Cert. Date: 07/02/2020
Solution File No.:	00915	Cert. No.: 00031
Sequential File No.:	00915	Lin. Date: 07/02/2020
		Lin. No.: 00031
		Soln. Date: 09/22/2020
		Soln. No.: 00246
		File Date: 09/22/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	19300	Expires: 10/30/2021
		Bottle No.: 1424

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:05D		
Control 1 EC	0.100%	02:06D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:06D		
Control 2 EC	0.099%	02:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:08D		
Control 3 EC	0.099%	02:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____

Badge No.: 331

Date: 09/22/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE	Calib. Date:	07/02/2020
Calibration File No.:	00905	Calib. No.:	00037
Certification File No.:	00906	Cert. Date:	07/02/2020
Linearity File No.:	00907	Lin. Date:	07/02/2020
Solution File No.:	00914	Soln. Date:	08/29/2020
Sequential File No.:	00914	File Date:	08/29/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0135
Solution Control Lot:	19300	Expires:	10/30/2021
		Bottle No.:	0082

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:37D		
Control 1 EC	0.100%	01:37D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:38D		
Control 2 EC	0.099%	01:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	01:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:40D		
Control 3 EC	0.098%	01:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____

Badge No.: 331

Date: 08/29/2020

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00905	Calib. Date: 07/02/2020
Certification File No.:	00906	Calib. No.: 00037
Linearity File No.:	00907	Cert. Date: 07/02/2020
Solution File No.:	00913	Cert. No.: 00031
Sequential File No.:	00913	Lin. Date: 07/02/2020
		Lin. No.: 00031
		Soln. Date: 08/28/2020
		Soln. No.: 00244
		File Date: 08/28/2020

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	19300	Expires: 10/30/2021
		Bottle No.: 1404

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:54D		
Control 1 EC	-.---%	23:55D	22.3°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	23:55D	22.3°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:		
Last Name: TIBBITT	First Name: DEAN	MI: M
Signature: 		Badge No.: 331
		Date: 08/29/2020

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00905	Calib. Date: 07/02/2020
Certification File No.:	00906	Calib. No.: 00037
Linearity File No.:	00907	Cert. Date: 07/02/2020
Solution File No.:	00911	Cert. No.: 00031
Sequential File No.:	00911	Lin. Date: 07/02/2020
		Lin. No.: 00031
		Soln. Date: 08/01/2020
		Soln. No.: 00243
		File Date: 08/01/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	19300	Expires: 10/30/2021
		Bottle No.: 1383

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:29D		
Control 1 EC	0.099%	19:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	19:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:31D		
Control 2 EC	0.097%	19:31D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:32D		
Control 3 EC	0.097%	19:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



0331

Badge No.: 331

Date: 08/01/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00893	Calib. Date: 01/14/2020
Certification File No.:	00894	Calib. No.: 00036
Linearity File No.:	00895	Cert. Date: 01/14/2020
Solution File No.:	00903	Cert. No.: 00030
Sequential File No.:	00903	Lin. Date: 01/14/2020
		Lin. No.: 00030
		Soln. Date: 06/02/2020
		Soln. No.: 00241
		File Date: 06/02/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	19300	Expires: 10/30/2021
		Bottle No.: 1399

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:15D		
Control 1 EC	0.100%	03:16D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:17D		
Control 2 EC	0.098%	03:17D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:18D		
Control 3 EC	0.098%	03:19D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



9-331

Badge No.: 331

Date: 06/02/2020

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00893	Calib. Date: 01/14/2020
Certification File No.:	00894	Calib. No.: 00036
Linearity File No.:	00895	Cert. Date: 01/14/2020
Solution File No.:	00901	Cert. No.: 00030
Sequential File No.:	00901	Lin. Date: 01/14/2020
		Lin. No.: 00030
		Soln. Date: 05/04/2020
		Soln. No.: 00240
		File Date: 05/04/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	19300	Expires: 10/30/2021
		Bottle No.: 1425

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:59D		
Control 1 EC	0.098%	14:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:01D		
Control 2 EC	0.097%	14:02D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	14:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02D		
Control 3 EC	0.097%	14:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____

Badge No.: 331

Date: 05/04/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00893	Calib. Date: 01/14/2020
Certification File No.:	00894	Calib. No.: 00036
Linearity File No.:	00895	Cert. Date: 01/14/2020
Solution File No.:	00900	Cert. No.: 00030
Sequential File No.:	00900	Lin. Date: 01/14/2020
		Lin. No.: 00030
		Soln. Date: 04/06/2020
		Soln. No.: 00239
		File Date: 04/06/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	19300	Expires: 10/31/2021
		Bottle No.: 1422

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:00D		
Control 1 EC	0.096%	22:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:01D		
Control 2 EC	0.096%	22:02D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:03D		
Control 3 EC	0.096%	22:03D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	22:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 04/06/2020

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00893	Calib. Date: 01/14/2020
Certification File No.:	00894	Calib. No.: 00036
Linearity File No.:	00895	Cert. Date: 01/14/2020
Solution File No.:	00899	Cert. No.: 00030
Sequential File No.:	00899	Lin. Date: 01/14/2020
		Lin. No.: 00030
		Soln. Date: 03/11/2020
		Soln. No.: 00238
		File Date: 03/11/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	19300	Expires: 10/30/2021
		Bottle No.: 1419

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:10S		
Control 1 EC	0.099%	14:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:11S		
Control 2 EC	0.097%	14:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:13S		
Control 3 EC	0.097%	14:13S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 03/11/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00893	Calib. Date: 01/14/2020
Certification File No.:	00894	Calib. No.: 00036
Linearity File No.:	00895	Cert. Date: 01/14/2020
Solution File No.:	00898	Cert. No.: 00030
Sequential File No.:	00898	Lin. Date: 01/14/2020
		Lin. No.: 00030
		Soln. Date: 02/07/2020
		Soln. No.: 00237
		File Date: 02/07/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	19300	Expires: 10/30/2021
		Bottle No.: 1396

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:40S		
Control 1 EC	0.099%	14:41S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:42S		
Control 2 EC	0.098%	14:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:43S		
Control 3 EC	0.098%	14:44S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 02/07/2020