

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00931	Calib. Date: 06/23/2021
Certification File No.:	00932	Calib. No.: 00039
Linearity File No.:	00933	Cert. Date: 06/23/2021
Solution File No.:	00946	Cert. No.: 00033
Sequential File No.:	00946	Lin. Date: 06/23/2021
		Lin. No.: 00033
		Soln. Date: 11/19/2021
		Soln. No.: 00260
		File Date: 11/19/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0659

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:34S		
Control 1 EC	0.100%	22:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:35S		
Control 2 EC	0.098%	22:36S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:37S		
Control 3 EC	0.098%	22:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:38S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature:  331

Badge No.: 331

Date: 11/19/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00931	Calib. Date:	06/23/2021
Certification File No.:	00932	Cert. No.:	00039
Linearity File No.:	00933	Cert. Date:	06/23/2021
Solution File No.:	00943	Lin. No.:	00033
Sequential File No.:	00943	Lin. Date:	06/23/2021
		Soln. Date:	10/19/2021
		Soln. No.:	00259
		File Date:	10/19/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0135
Solution Control Lot:	20390	Expires:	09/02/2022
		Bottle No.:	0093

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:04D		
Control 1 EC	0.101%	07:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:05D		
Control 2 EC	0.099%	07:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:06D		
Control 3 EC	0.099%	07:07D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 10/19/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00931	Calib. Date: 06/23/2021
Certification File No.:	00932	Calib. No.: 00039
Linearity File No.:	00933	Cert. Date: 06/23/2021
Solution File No.:	00942	Cert. No.: 00033
Sequential File No.:	00942	Lin. Date: 06/23/2021
		Lin. No.: 00033
		Soln. Date: 09/21/2021
		Soln. No.: 00258
		File Date: 09/21/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0034

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:14D		
Control 1 EC	0.098%	08:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16D		
Control 2 EC	0.097%	08:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	08:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17D		
Control 3 EC	0.097%	08:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBETT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 09/21/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00931	Calib. Date: 06/23/2021
Certification File No.:	00932	Calib. No.: 00039
Linearity File No.:	00933	Cert. Date: 06/23/2021
Solution File No.:	00940	Cert. No.: 00033
Sequential File No.:	00940	Lin. Date: 06/23/2021
		Lin. No.: 00033
		Soln. Date: 08/23/2021
		Soln. No.: 00257
		File Date: 08/23/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0146

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:04D		
Control 1 EC	0.100%	05:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:06D		
Control 2 EC	0.099%	05:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:07D		
Control 3 EC	0.099%	05:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 08/23/2021

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00931	Calib. Date: 06/23/2021
Certification File No.:	00932	Calib. No.: 00039
Linearity File No.:	00933	Cert. Date: 06/23/2021
Solution File No.:	00937	Cert. No.: 00033
Sequential File No.:	00937	Lin. Date: 06/23/2021
		Lin. No.: 00033
		Soln. Date: 07/28/2021
		Soln. No.: 00256
		File Date: 07/28/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0028

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:15D		
Control 1 EC	0.100%	17:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:16D		
Control 2 EC	0.099%	17:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:17D		
Control 3 EC	0.099%	17:18D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:18D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____

Badge No.: 331

Date: 07/28/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00919	Calib. Date: 12/23/2020
Certification File No.:	00920	Calib. No.: 00038
Linearity File No.:	00921	Cert. Date: 12/23/2020
Solution File No.:	00930	Cert. No.: 00032
Sequential File No.:	00930	Lin. Date: 12/23/2020
		Lin. No.: 00032
		Soln. Date: 05/28/2021
		Soln. No.: 00254
		File Date: 05/28/2021

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXC S3-0154
Control Solution %: 0.100%		Expires: 09/02/2022
Solution Control Lot: 20390		Bottle No.: 0082

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:02D		
Control 1 EC	0.100%	08:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:03D		
Control 2 EC	0.099%	08:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:05D		
Control 3 EC	0.099%	08:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



- 331

Badge No.: 331

Date: 05/28/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00919	Calib. Date: 12/23/2020
Certification File No.:	00920	Calib. No.: 00038
Linearity File No.:	00921	Cert. Date: 12/23/2020
Solution File No.:	00928	Cert. No.: 00032
Sequential File No.:	00928	Lin. Date: 12/23/2020
		Lin. No.: 00032
		Soln. Date: 05/01/2021
		Soln. No.: 00253
		File Date: 05/01/2021

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0091

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:45D		
Control 1 EC	0.100%	10:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46D		
Control 2 EC	0.099%	10:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		
Control 3 EC	0.098%	10:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:49D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: 

Badge No.: 331

Date: 05/01/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00919	Calib. Date: 12/23/2020
Certification File No.:	00920	Calib. No.: 00038
Linearity File No.:	00921	Cert. Date: 12/23/2020
Solution File No.:	00927	Cert. No.: 00032
Sequential File No.:	00927	Lin. Date: 12/23/2020
		Lin. No.: 00032
		Soln. Date: 03/29/2021
		Soln. No.: 00252
		File Date: 03/29/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0142

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:55S		
Control 1 EC	0.100%	14:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57S		
Control 2 EC	0.099%	14:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58S		
Control 3 EC	0.099%	14:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____

Badge No.: 331

Date: 03/29/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00919	Calib. Date: 12/23/2020
Certification File No.:	00920	Calib. No.: 00038
Linearity File No.:	00921	Cert. Date: 12/23/2020
Solution File No.:	00925	Cert. No.: 00032
Sequential File No.:	00925	Lin. Date: 12/23/2020
		Lin. No.: 00032
		Soln. Date: 02/17/2021
		Soln. No.: 00251
		File Date: 02/17/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0256

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:34S		
Control 1 EC	0.102%	19:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:35S		
Control 2 EC	0.098%	19:36S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	19:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:37S		
Control 3 EC	0.098%	19:37S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	19:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 02/17/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00919	Calib. Date: 12/23/2020
Certification File No.:	00920	Calib. No.: 00038
Linearity File No.:	00921	Cert. Date: 12/23/2020
Solution File No.:	00924	Cert. No.: 00032
Sequential File No.:	00924	Lin. Date: 12/23/2020
		Lin. No.: 00032
		Soln. Date: 01/20/2021
		Soln. No.: 00250
		File Date: 01/20/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0258

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:19S		
Control 1 EC	0.098%	00:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:20S		
Control 2 EC	0.097%	00:21S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:22S		
Control 3 EC	0.097%	00:22S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 01/20/2021