

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00989	Calib. Date: 07/19/2023
Certification File No.:	00990	Calib. No.: 00044
Linearity File No.:	00991	Cert. Date: 07/19/2023
Solution File No.:	01003	Cert. No.: 00038
Sequential File No.:	01003	Lin. Date: 07/19/2023
		Lin. No.: 00038
		Soln. Date: 12/06/2023
		Soln. No.: 00286
		File Date: 12/06/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	23110	Expires: 04/12/2025
		Bottle No.: 0624

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:32S		
Control 1 EC	0.100%	10:32S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:33S		
Control 2 EC	0.099%	10:34S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:35S		
Control 3 EC	0.099%	10:35S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:36S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SLATTERY

First Name: SHAUNE

MI: M

Signature: 

Badge No.: 334

Date: 12/06/2023

Calibrating Unit New Standard Solution Report


Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0067
Location: LONGPORT POLICE
Calibration File No.: 00989 Calib. Date: 07/19/2023 Calib. No.: 00044
Certification File No.: 00990 Cert. Date: 07/19/2023 Cert. No.: 00038
Linearity File No.: 00991 Lin. Date: 07/19/2023 Lin. No.: 00038
Solution File No.: 01000 Soln. Date: 12/02/2023 Soln. No.: 00285
Sequential File No.: 01000 File Date: 12/02/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0154
Control Solution %: 0,100% Expires: 04/12/2025
Solution Control Lot: 23110 Bottle No.: 0659

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:48S		
Control 1 EC	0.100%	18:49S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:49S		
Control 2 EC	0.098%	18:50S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:51S		
Control 3 EC	0.098%	18:52S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: TIBBITT First Name: DEAN MI: M
Signature:  Badge No.: 331
Date: 12/02/2023

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00989	Calib. Date:	07/19/2023
Certification File No.:	00990	Cert. No.:	00044
Linearity File No.:	00991	Cert. Date:	07/19/2023
Solution File No.:	00999	Lin. No.:	00038
Sequential File No.:	00999	Soln. Date:	11/06/2023
		File Date:	11/06/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0154
Solution Control Lot:	23110	Expires:	04/12/2025
		Bottle No.:	0410

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:12S		
Control 1 EC	0.099%	08:12S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:13S		
Control 2 EC	0.099%	08:14S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:15S		
Control 3 EC	0.098%	08:15S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:15S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____

Badge No.: 331

Date: 11/06/2023

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	00989	Calib. Date:	07/19/2023
Certification File No.:	00990	Cert. No.:	00044
Linearity File No.:	00991	Cert. Date:	07/19/2023
Solution File No.:	00997	Lin. Date:	07/19/2023
Sequential File No.:	00997	Soln. Date:	10/05/2023
		File Date:	10/05/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0154
Solution Control Lot:	23110	Expires:	04/12/2025
		Bottle No.:	0656

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:53D		
Control 1 EC	0.101%	05:54D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:54D		
Control 2 EC	0.100%	05:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	05:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:56D		
Control 3 EC	0.100%	05:57D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: TIBBATT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 10/05/2023

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0067
Location: LONGPORT POLICE
Calibration File No.: 00989 Calib. Date: 07/19/2023 Calib. No.: 00044
Certification File No.: 00990 Cert. Date: 07/19/2023 Cert. No.: 00038
Linearity File No.: 00991 Lin. Date: 07/19/2023 Lin. No.: 00038
Solution File No.: 00994 Soln. Date: 09/08/2023 Soln. No.: 00282
Sequential File No.: 00994 File Date: 09/08/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0154
Control Solution %: 0.100% Expires: 04/12/2025
Solution Control Lot: 23110 Bottle No.: 0660

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:10D		
Control 1 EC	0.103%	12:11D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:12D		
Control 2 EC	0.101%	12:13D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13D		
Control 3 EC	0.101%	12:14D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: 

Badge No.: 331

Date: 09/08/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00989	Calib. Date: 07/19/2023
Certification File No.:	00990	Calib. No.: 00044
Linearity File No.:	00991	Cert. Date: 07/19/2023
Solution File No.:	00993	Cert. No.: 00038
Sequential File No.:	00993	Lin. Date: 07/19/2023
		Lin. No.: 00038
		Soln. Date: 08/17/2023
		Soln. No.: 00281
		File Date: 08/17/2023

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXC S3-0154
Control Solution %:	0.100%		Expires: 04/12/2025
Solution Control Lot:	23110		Bottle No.: 0414

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:20D		
Control 1 EC	0.102%	01:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	01:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:21D		
Control 2 EC	0.101%	01:22D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	01:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:22D		
Control 3 EC	0.101%	01:23D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 08/17/2023



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor
SHEILA Y. OLIVER
Lt. Governor

MATTHEW J. PLATKIN
Acting Attorney General
COLONEL PATRICK J. CALLAHAN
Superintendent

CERTIFICATION OF ANALYSIS
0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 05/15/2023

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 23110

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1204 to 0.1215 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is April 12, 2025.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Michael Kennedy
OFS Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 17 day of May, 2023.
Notary

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522
My Commission Expires 8/13/2024



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00979	Calib. Date: 03/01/2023
Certification File No.:	00980	Calib. No.: 00043
Linearity File No.:	00981	Cert. Date: 03/01/2023
Solution File No.:	00986	Cert. No.: 00037
Sequential File No.:	00986	Lin. Date: 03/01/2023
		Lin. No.: 00037
		Soln. Date: 06/22/2023
		Soln. No.: 00279
		File Date: 06/22/2023

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXC S3-0154
Control Solution %:	0.100%		Expires: 02/17/2024
Solution Control Lot:	22070		Bottle No.: 0540

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:53D		
Control 1 EC	0.102%	05:54D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:55D		
Control 2 EC	0.099%	05:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	05:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:56D		
Control 3 EC	0.099%	05:57D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	05:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:57D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature:  331

Badge No.: 331

Date: 06/22/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00979	Calib. Date: 03/01/2023
Certification File No.:	00980	Calib. No.: 00043
Linearity File No.:	00981	Cert. Date: 03/01/2023
Solution File No.:	00985	Cert. No.: 00037
Sequential File No.:	00985	Lin. Date: 03/01/2023
		Lin. No.: 00037
		Soln. Date: 05/16/2023
		Soln. No.: 00278
		File Date: 05/16/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0682

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:37D		
Control 1 EC	0.101%	07:38D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:39D		
Control 2 EC	0.098%	07:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	07:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40D		
Control 3 EC	0.098%	07:41D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	07:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:41D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SLATTERY

First Name: SHAUNE

MI: M

Signature:  #334

Badge No.: 334

Date: 05/16/2023

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0067
Location: LONGPORT POLICE
Calibration File No.: 00979 Calib. Date: 03/01/2023 Calib. No.: 00043
Certification File No.: 00980 Cert. Date: 03/01/2023 Cert. No.: 00037
Linearity File No.: 00981 Lin. Date: 03/01/2023 Lin. No.: 00037
Solution File No.: 00983 Soln. Date: 03/12/2023 Soln. No.: 00276
Sequential File No.: 00983 File Date: 03/12/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0154
Control Solution %: 0.100% Expires: 07/21/2023
Solution Control Lot: 21270 Bottle No.: 0160

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:15S		
Control 1 EC	0.101%	19:16S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:16S		
Control 2 EC	0.099%	19:17S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:18S		
Control 3 EC	0.100%	19:19S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SLATTERY

First Name: SHAUNE

MI: M

Signature: SM Slattery #334

Badge No.: 334

Date: 03/12/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00970	Calib. Date: 09/29/2022
Certification File No.:	00971	Calib. No.: 00042
Linearity File No.:	00972	Cert. Date: 09/29/2022
Solution File No.:	00978	Cert. No.: 00036
Sequential File No.:	00978	Lin. Date: 09/29/2022
		Lin. No.: 00036
		Soln. Date: 02/13/2023
		Soln. No.: 00274
		File Date: 02/13/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0161

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:33S		
Control 1 EC	0.099%	17:34S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:35S		
Control 2 EC	0.098%	17:35S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	17:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:36S		
Control 3 EC	0.098%	17:37S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SLATTERY

First Name: SHAUNE

MI: M

Signature: _____

SM Slattery #334

Badge No.: 334

Date: 02/13/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00970	Calib. Date: 09/29/2022
Certification File No.:	00971	Calib. No.: 00042
Linearity File No.:	00972	Cert. Date: 09/29/2022
Solution File No.:	00977	Cert. No.: 00036
Sequential File No.:	00977	Lin. Date: 09/29/2022
		Lin. No.: 00036
		Soln. Date: 01/05/2023
		Soln. No.: 00273
		File Date: 01/05/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0135
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0329

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:25S		
Control 1 EC	0.100%	15:25S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:26S		
Control 2 EC	0.099%	15:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:27S		
Control 3 EC	0.100%	15:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:29S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SLATTERY

First Name: SHAUNE

MI: M

Signature:  # 334

Badge No.: 334

Date: 01/05/2023

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0067
Location: LONGPORT POLICE
Calibration File No.: 00970 Calib. Date: 09/29/2022 Calib. No.: 00042
Certification File No.: 00971 Cert. Date: 09/29/2022 Cert. No.: 00036
Linearity File No.: 00972 Lin. Date: 09/29/2022 Lin. No.: 00036
Solution File No.: 00976 Soln. Date: 01/04/2023 Soln. No.: 00272
Sequential File No.: 00976 File Date: 01/04/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135
Control Solution %: 0.100% Expires: 07/21/2023
Solution Control Lot: 21270 Bottle No.: 0622

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:46S		
Control 1 EC	0.101%	08:47S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48S		
Control 2 EC	0.099%	08:48S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:48S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:49S		
Control 3 EC	0.100%	08:50S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SLATTERY

First Name: SHAUNE

MI: M

Signature: SM Slattery #334

Badge No.: 334

Date: 01/04/2023