

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0067
Location: LONGPORT POLICE
Calibration File No.: 01012 Calib. Date: 05/13/2024 Calib. No.: 00046
Certification File No.: 01013 Cert. Date: 05/13/2024 Cert. No.: 00040
Linearity File No.: 01014 Lin. Date: 05/13/2024 Lin. No.: 00040
Solution File No.: 01018 Soln. Date: 07/16/2024 Soln. No.: 00293
Sequential File No.: 01018 File Date: 07/16/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135
Control Solution %: 0.100% Expires: 01/09/2026
Solution Control Lot: 24020 Bottle No.: 0328

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:55D		
Control 1 EC	0.101%	23:56D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:57D		
Control 2 EC	0.100%	23:57D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:58D		
Control 3 EC	0.099%	23:59D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____

Badge No.: 331

Date: 07/17/2024

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0067
Location: LONGPORT POLICE
Calibration File No.: 01012 **Calib. Date:** 05/13/2024 **Calib. No.:** 00046
Certification File No.: 01013 **Cert. Date:** 05/13/2024 **Cert. No.:** 00040
Linearity File No.: 01014 **Lin. Date:** 05/13/2024 **Lin. No.:** 00040
Solution File No.: 01017 **Soln. Date:** 06/15/2024 **Soln. No.:** 00292
Sequential File No.: 01017 **File Date:** 06/15/2024

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXC S3-0135
Control Solution %: 0.100% **Expires:** 01/09/2026
Solution Control Lot: 24020 **Bottle No.:** 0451

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:49D		
Control 1 EC	0.100%	01:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:50D		
Control 2 EC	0.099%	01:51D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:52D		
Control 3 EC	0.098%	01:52D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:53D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: 

Badge No.: 331

Date: 06/15/2024

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0067
Location:	LONGPORT POLICE		
Calibration File No.:	01004	Calib. Date:	12/20/2023
Certification File No.:	01005	Cert. No.:	00045
Linearity File No.:	01006	Cert. Date:	12/20/2023
Solution File No.:	01011	Lin. No.:	00039
Sequential File No.:	01011	Lin. Date:	12/20/2023
		Soln. Date:	04/25/2024
		Soln. No.:	00290
		File Date:	04/25/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0135
Solution Control Lot:	23110	Expires:	04/12/2025
		Bottle No.:	0517

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:52D		
Control 1 EC	0.100%	19:52D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:53D		
Control 2 EC	0.099%	19:54D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:55D		
Control 3 EC	0.099%	19:55D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____



Badge No.: 331

Date: 04/25/2024

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0067
Location: LONGPORT POLICE
Calibration File No.: 01004 Calib. Date: 12/20/2023 Calib. No.: 00045
Certification File No.: 01005 Cert. Date: 12/20/2023 Cert. No.: 00039
Linearity File No.: 01006 Lin. Date: 12/20/2023 Lin. No.: 00039
Solution File No.: 01009 Soln. Date: 03/26/2024 Soln. No.: 00289
Sequential File No.: 01009 File Date: 03/26/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135
Control Solution %: 0.100% Expires: 04/12/2025
Solution Control Lot: 23110 Bottle No.: 0578

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:13S		
Control 1 EC	0.100%	06:13S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:13S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:14S		
Control 2 EC	0.098%	06:15S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.098%	06:15S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:15S		
Control 3 EC	0.098%	06:16S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:16S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:17S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: _____

Badge No.: 331

Date: 03/26/2024

Calibrating Unit New Standard Solution Report


Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0067
Location: LONGPORT POLICE
Calibration File No.: 01004 Calib. Date: 12/20/2023 Calib. No.: 00045
Certification File No.: 01005 Cert. Date: 12/20/2023 Cert. No.: 00039
Linearity File No.: 01006 Lin. Date: 12/20/2023 Lin. No.: 00039
Solution File No.: 01008 Soln. Date: 02/23/2024 Soln. No.: 00288
Sequential File No.: 01008 File Date: 02/23/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135
Control Solution %: 0.100% Expires: 04/12/2025
Solution Control Lot: 23110 Bottle No.: 0654

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:06S		
Control 1 EC	0.100%	02:06S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:06S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:07S		
Control 2 EC	0.097%	02:08S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:08S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:09S		
Control 3 EC	0.098%	02:09S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:09S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: TIBBITT First Name: DEAN MI: M
Signature:  Badge No.: 331
Date: 02/23/2024