

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
Location:	BUENA BOROUGH POLICE		
Calibration File No.:	00243	Calib. Date:	12/04/2013
Certification File No.:	00244	Cert. Date:	12/04/2013
Linearity File No.:	00245	Lin. Date:	12/04/2013
Solution File No.:	00254	Soln. Date:	02/01/2014
Sequential File No.:	00254	File Date:	02/01/2014

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXJ S3-0210
Control Solution %:	0.100%			Expires:	07/18/2015
Solution Control Lot:	13G115			Bottle No.:	0928

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:42S		
Control 1 EC	0.100%	08:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:43S		
Control 2 EC	0.100%	08:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:44S		
Control 3 EC	0.099%	08:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEUDLE

First Name: DENNIS

MI: W

Signature: _____



Badge No.: 1159

Date: 02/01/2014

Dräger

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe
DDWJ P2-204

Certification date:
4-29-13

Next Certification due:
4-29-14

Probe Value
106

Dräger Safety Diagnostics, Inc.
Technical Service Department

AC



Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDXJ 53-0210

Certification Date

4-29-13

Technician

AC

Re-Certification Due Date

4-29-14