

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
<b>Location:</b>	BUENA BOROUGH POLICE		
<b>Calibration File No.:</b>	00140	<b>Calib. Date:</b>	08/15/2011
<b>Certification File No.:</b>	00141	<b>Cert. No.:</b>	00007
<b>Linearity File No.:</b>	00142	<b>Lin. Date:</b>	08/15/2011
<b>Solution File No.:</b>	00144	<b>Soln. No.:</b>	00045
<b>Sequential File No.:</b>	00144	<b>File Date:</b>	09/08/2011
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b>	DDXJ S3-0210
<b>Solution Control Lot:</b>	09L072	<b>Expires:</b>	12/31/2011
		<b>Bottle No.:</b>	0937

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:14D		
Control 1 EC	0.100%	11:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15D		
Control 2 EC	0.098%	11:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17D		
Control 3 EC	0.099%	11:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature: 

Badge No.: 1106

Date: 09/08/2011

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
<b>Location:</b>	BUENA BOROUGH POLICE	
<b>Calibration File No.:</b>	00105	<b>Calib. Date:</b> 02/15/2011
<b>Certification File No.:</b>	00106	<b>Cert. No.:</b> 00006
<b>Linearity File No.:</b>	00107	<b>Lin. Date:</b> 02/15/2011
<b>Solution File No.:</b>	00111	<b>Lin. No.:</b> 00006
<b>Sequential File No.:</b>	00111	<b>Soln. Date:</b> 03/21/2011
		<b>Soln. No.:</b> 00038
		<b>File Date:</b> 03/21/2011
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b> CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b> DDBJ-0006
<b>Solution Control Lot:</b>	09L072	<b>Expires:</b> 12/31/2011
		<b>Bottle No.:</b> 0928

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:26S		
Control 1 EC	0.099%	10:27S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:27S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:28S		
Control 2 EC	0.098%	10:28S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:28S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:29S		
Control 3 EC	0.097%	10:30S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:30S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature: \_\_\_\_\_



Badge No.: 1106

Date: 03/21/2011

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXB-0074  
**Location:** BUENA BOROUGH POLICE  
**Calibration File No.:** 00086 **Calib. Date:** 08/31/2010 **Calib. No.:** 00010  
**Certification File No.:** 00087 **Cert. Date:** 08/31/2010 **Cert. No.:** 00005  
**Linearity File No.:** 00088 **Lin. Date:** 08/31/2010 **Lin. No.:** 00005  
**Solution File No.:** 00099 **Soln. Date:** 12/09/2010 **Soln. No.:** 00034  
**Sequential File No.:** 00099 **File Date:** 12/09/2010

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDBJ-0006  
**Control Solution %:** 0.100% **Expires:** 12/31/2011  
**Solution Control Lot:** 09L072 **Bottle No.:** 0949

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:15S		
Control 1 EC	0.100%	16:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:16S		
Control 2 EC	0.099%	16:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:18S		
Control 3 EC	0.098%	16:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature: 

Badge No.: 1106

Date: 12/09/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
<b>Location:</b>	BUENA BOROUGH POLICE	
<b>Calibration File No.:</b>	00086	<b>Calib. Date:</b> 08/31/2010
<b>Certification File No.:</b>	00087	<b>Calib. No.:</b> 00010
<b>Linearity File No.:</b>	00088	<b>Cert. Date:</b> 08/31/2010
<b>Solution File No.:</b>	00094	<b>Cert. No.:</b> 00005
<b>Sequential File No.:</b>	00094	<b>Lin. Date:</b> 08/31/2010
		<b>Lin. No.:</b> 00005
		<b>Soln. Date:</b> 10/27/2010
		<b>Soln. No.:</b> 00033
		<b>File Date:</b> 10/27/2010
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b> CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b> DDBJ-0006
<b>Solution Control Lot:</b>	09L072	<b>Expires:</b> 12/31/2011
		<b>Bottle No.:</b> 0929

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:00D		
Control 1 EC	0.100%	14:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:01D		
Control 2 EC	0.100%	14:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02D		
Control 3 EC	0.100%	14:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature:



Badge No.: 1106

Date: 10/27/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
Location:	BUENA BOROUGH POLICE	
Calibration File No.:	00086	Calib. Date: 08/31/2010
Certification File No.:	00087	Calib. No.: 00010
Linearity File No.:	00088	Cert. Date: 08/31/2010
Solution File No.:	00089	Cert. No.: 00005
Sequential File No.:	00089	Lin. Date: 08/31/2010
		Lin. No.: 00005
		Soln. Date: 09/24/2010
		Soln. No.: 00032
		File Date: 09/24/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDBJ-0006
Solution Control Lot:	09L072	Expires: 12/31/2011
		Bottle No.: 0934

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:51D		
Control 1 EC	0.100%	12:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:53D		
Control 2 EC	0.099%	12:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:54D		
Control 3 EC	0.099%	12:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDWTPD-088 J<sup>th</sup>

**Changed By:**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: Joseph Dellanoce # 6027

Badge No.: 6027

Date: 09/24/2010

# Calibrating Unit New Standard Solution Report

**Equipment**  
 Location:  
 Calibration File No.:  
 Certification File No.:  
 Linearity File No.:  
 Solution File No.:  
 Sequential File No.:

Alcotest 7110 MKIII-C  
 BUENA BOROUGH POLICE

Calib. Date: 03/30/2010  
 Cert. Date: 03/30/2010  
 Lin. Date: 03/30/2010  
 Soln. Date: 07/13/2010  
 File Date: 07/13/2010

Serial No.: ARXB-0074

Calib. No.: 00005  
 Cert. No.: 00004  
 Lin. No.: 00004  
 Soln. No.: 00025

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 09L072

Model No.: CU-34

Serial No.: DDXJ S3-0209  
 Expires: 12/31/2011  
 Bottle No.: 0940

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:20D		*** TEST PASSED ***
Control 1 EC	0.099%	14:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:21D	34.0°C	
Ambient Air Blank	0.000%	14:21D		*** TEST PASSED ***
Control 2 EC	0.098%	14:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:22D	34.0°C	
Ambient Air Blank	0.000%	14:23D		*** TEST PASSED ***
Control 3 EC	0.098%	14:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:23D	34.0°C	
Ambient Air Blank	0.000%	14:24D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RETRILLO

First Name: LAWRENCE

MI: A

Signature: 

Badge No.: 1106

Date: 07/13/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
Location:	BUENA BOROUGH POLICE	Calib. No.:	00005
Calibration File No.:	00055	Cert. No.:	00004
Certification File No.:	00056	Lin. No.:	00004
Linearity File No.:	00057	Soln. No.:	00024
Solution File No.:	00066	Calib. Date:	03/30/2010
Sequential File No.:	00066	Cert. Date:	03/30/2010
		Lin. Date:	03/30/2010
		Soln. Date:	06/01/2010
		File Date:	06/01/2010
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXJ S3-0209
Solution Control Lot:	09L072	Expires:	12/31/2011
		Bottle No.:	0935

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:12D		
Control 1 EC	0.100%	13:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13D		
Control 2 EC	0.099%	13:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14D		
Control 3 EC	0.098%	13:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature:



Badge No.: 1106

Date: 06/01/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
<b>Location:</b>	BUENA BOROUGH POLICE		
<b>Calibration File No.:</b>	00055	<b>Calib. Date:</b>	03/30/2010
<b>Certification File No.:</b>	00056	<b>Cert. No.:</b>	00005
<b>Linearity File No.:</b>	00057	<b>Cert. Date:</b>	03/30/2010
<b>Solution File No.:</b>	00060	<b>Lin. Date:</b>	03/30/2010
<b>Sequential File No.:</b>	00060	<b>Soln. Date:</b>	04/30/2010
		<b>File Date:</b>	04/30/2010
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b>	DDXJ S3-0209
<b>Solution Control Lot:</b>	09L072	<b>Expires:</b>	12/31/2011
		<b>Bottle No.:</b>	0939

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:56D		
Control 1 EC	0.101%	14:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58D		
Control 2 EC	0.099%	14:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:59D		
Control 3 EC	0.098%	14:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature: 

Badge No.: 1106

Date: 04/30/2010



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
Location:	BUENA BOROUGH POLICE	Calib. No.:	00004
Calibration File No.:	00039	Cert. Date:	10/05/2009
Certification File No.:	00040	Lin. Date:	10/05/2009
Linearity File No.:	00041	Soln. Date:	02/18/2010
Solution File No.:	00050	File Date:	02/18/2010
Sequential File No.:	00050		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXJ S3-0209
Solution Control Lot:	09I069	Expires:	09/18/2011
		Bottle No.:	0944

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:44S		
Control 1 EC	0.102%	12:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:45S		
Control 2 EC	0.099%	12:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:47S		
Control 3 EC	0.099%	12:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature: 

Badge No.: 1106

Date: 02/18/2010



State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
POST OFFICE BOX 7068  
WEST TRENTON NJ 08628-0068  
(609) 882-2000

JON S. CORZINE  
Governor

ANNE MILGRAM  
Attorney General

COLONEL JOSEPH R. FUENTES  
Superintendent

**CERTIFICATION OF ANALYSIS  
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

**ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION:** Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

**MANUFACTURER:** Drager Safety, Inc.      **ANALYSIS DATE:** 10/13/2009

**BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER:** 091069

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1199 to 0.1204 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is September 18, 2011.

As Forensic Laboratory Director of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Howard J. Baum, Ph.D.  
Forensic Laboratory Director  
Division of State Police

Sworn to and subscribed before me this 30<sup>th</sup> day of October, 2009.

Notary

Linda L. Demerle  
Notary Public, New Jersey  
My Commission Expires 6-17-14



New Jersey Is An Equal Opportunity Employer  
Printed on Recycled Paper and Recyclable



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
<b>Location:</b>	BUENA BOROUGH POLICE	
<b>Calibration File No.:</b>	00039	Calib. Date: 10/05/2009
<b>Certification File No.:</b>	00040	Cert. Date: 10/05/2009
<b>Linearity File No.:</b>	00041	Lin. Date: 10/05/2009
<b>Solution File No.:</b>	00049	Soln. Date: 12/31/2009
<b>Sequential File No.:</b>	00049	File Date: 12/31/2009
<b>Calibrating Unit:</b>	WET	Model No.: CU-34
<b>Control Solution %:</b>	0.100%	Serial No.: DDXJ S3-0209
<b>Solution Control Lot:</b>	08B050	Expires: 02/13/2010
		Bottle No.: 0484

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:59S		*** TEST PASSED ***
Control 1 EC	0.101%	19:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:00S		*** TEST PASSED ***
Control 2 EC	0.098%	20:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:01S		*** TEST PASSED ***
Control 3 EC	0.100%	20:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:03S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature: 

Badge No.: 1106

Date: 12/31/2009



Drägersafety

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

**Dräger Safety Diagnostics, Inc.**

Model: ALCOTEST® CU34

Model: MARK IIA

Other: \_\_\_\_\_

Serial Number:

DDXJS3-0209

Certification Date

Technician

Re-Certification Due Date

10-5-09

ZM

10-5-10

## Drägersafety

### ALCOTEST® 7110 TEMPERATURE PROBE

# CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

Certification date:

Next Certification due:

DDWA 12-148

10-5-09

10-5-10

Probe Value

104

Dräger Safety Diagnostics, Inc.  
Technical Service Department

ZM

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
Location:	BUENA BOROUGH POLICE	Calib. No.:	00004
Calibration File No.:	00039	Cert. No.:	00003
Certification File No.:	00040	Lin. No.:	00003
Linearity File No.:	00041	Soln. No.:	00019
Solution File No.:	00047		
Sequential File No.:	00047		
		Calib. Date:	10/05/2009
		Cert. Date:	10/05/2009
		Lin. Date:	10/05/2009
		Soln. Date:	11/14/2009
		File Date:	11/14/2009
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXJ S3-0209
Solution Control Lot:	08B050	Expires:	02/13/2010
		Bottle No.:	0496

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:28S		*** TEST PASSED ***
Control 1 EC	0.099%	13:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:29S		*** TEST PASSED ***
Control 2 EC	0.097%	13:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:31S		*** TEST PASSED ***
Control 3 EC	0.097%	13:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:32S		*** TEST PASSED ***

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature:

Badge No.: 1106

Date: 11/14/2009

## Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
Location:	BUENA BOROUGH POLICE	
Calibration File No.:	00039	Calib. Date: 10/05/2009
Certification File No.:	00040	Cert. No.: 00004
Linearity File No.:	00041	Cert. Date: 10/05/2009
Solution File No.:	00043	Lin. Date: 10/05/2009
Sequential File No.:	00043	Lin. No.: 00003
		Soln. Date: 10/15/2009
		Soln. No.: 00018
		File Date: 10/15/2009
<b>Calibrating Unit:</b>	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXJ S3-0209
Solution Control Lot:	08B050	Expires: 02/13/2010
		Bottle No.: 0497

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:23D		
Control 1 EC	0.100%	15:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25D		
Control 2 EC	0.098%	15:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:26D		
Control 3 EC	0.098%	15:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: *DPWAP2-148*

**Changed By:**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: *J. Dellanoce*

Badge No.: 6027

Date: 10/15/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
Location:	BUENA BOROUGH POLICE		
Calibration File No.:	00022	Calib. Date:	04/24/2009
Certification File No.:	00023	Calib. No.:	00003
Linearity File No.:	00024	Cert. Date:	04/24/2009
Solution File No.:	00037	Cert. No.:	00002
Sequential File No.:	00037	Lin. Date:	04/24/2009
		Lin. No.:	00002
		Soln. Date:	09/03/2009
		Soln. No.:	00016
		File Date:	09/03/2009
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXJ S3-0210
Solution Control Lot:	08B050	Expires:	02/13/2010
		Bottle No.:	0492

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:56D		
Control 1 EC	0.101%	18:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:57D		
Control 2 EC	0.099%	18:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:59D		
Control 3 EC	0.100%	18:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: **PETRILLO**

First Name: **LAWRENCE**

MI: **A**

Signature:

Badge No.: 1106

Date: 09/03/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
Location:	BUENA BOROUGH POLICE	Calib. No.:	00003
Calibration File No.:	00022	Cert. Date:	04/24/2009
Certification File No.:	00023	Lin. Date:	04/24/2009
Linearity File No.:	00024	Soln. Date:	08/12/2009
Solution File No.:	00035	File Date:	08/12/2009
Sequential File No.:	00035		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXJ S3-0210
Solution Control Lot:	08B050	Expires:	02/13/2010
		Bottle No.:	0479

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:41D		
Control 1 EC	0.100%	22:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:42D		
Control 2 EC	0.100%	22:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:44D		
Control 3 EC	0.099%	22:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	22:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**  
 Last Name: PETRILLO      First Name: LAWRENCE      MI: A  
 Signature:       Badge No.: 1106  
 Date: 08/12/2009



# Calibrating Unit New Standard Solution Report

## Equipment

Location: BUENA BOROUGH POLICE  
 Calibration File No.: 00022  
 Certification File No.: 00023  
 Linearity File No.: 00024  
 Solution File No.: 00034  
 Sequential File No.: 00034

Alcotest 7110 MKIII-C  
 BUENA BOROUGH POLICE

Calib. Date: 04/24/2009  
 Cert. Date: 04/24/2009  
 Lin. Date: 04/24/2009  
 Soln. Date: 07/11/2009  
 File Date: 07/11/2009

Serial No.: ARXB-0074

Calib. No.: 00003  
 Cert. No.: 00002  
 Lin. No.: 00002  
 Soln. No.: 00014

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 08B050

Model No.: CU-34

Serial No.: DDXJ S3-0210  
 Expires: 02/13/2010  
 Bottle No.: 0487

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:59D		
Control 1 EC	0.101%	16:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:00D		
Control 2 EC	0.099%	17:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:01D		
Control 3 EC	0.100%	17:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature:



Badge No.: 1106

Date: 07/11/2009

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C  
**Location:** BUENA BOROUGH POLICE  
**Calibration File No.:** 00022 **Calib. Date:** 04/24/2009 **Calib. No.:** 00003  
**Certification File No.:** 00023 **Cert. Date:** 04/24/2009 **Cert. No.:** 00002  
**Linearity File No.:** 00024 **Lin. Date:** 04/24/2009 **Lin. No.:** 00002  
**Solution File No.:** 00032 **Soln. Date:** 06/10/2009 **Soln. No.:** 00013  
**Sequential File No.:** 00032 **File Date:** 06/10/2009

Serial No.: ARXB-0074

**Calibrating Unit:** WET  
**Control Solution %:** 0.100%  
**Solution Control Lot:** 08B050

Model No.: CU-34

Serial No.: DDXJ S3-0210  
**Expires:** 02/13/2010  
**Bottle No.:** 0489

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:57D		
Control 1 EC	0.100%	19:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:59D		
Control 2 EC	0.100%	19:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:00D		
Control 3 EC	0.099%	20:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature: 

Badge No.: 1106

Date: 06/10/2009

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
<b>Location:</b>	BUENA BOROUGH POLICE	
<b>Calibration File No.:</b>	00004	Calib. Date: 11/05/2008
<b>Certification File No.:</b>	00005	Cert. No.: 00002
<b>Linearity File No.:</b>	00006	Cert. Date: 11/05/2008
<b>Solution File No.:</b>	00021	Lin. No.: 00001
<b>Sequential File No.:</b>	00021	Soln. Date: 03/30/2009
		Soln. No.: 00009
		File Date: 03/30/2009
<b>Calibrating Unit:</b>	WET	Model No.: CU-34
<b>Control Solution %:</b>	0.100%	Serial No.: DDXJ S3-0210
<b>Solution Control Lot:</b>	08B050	Expires: 02/13/2010
		Bottle No.: 0482

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:04S	09:04 DJS	
Control 1 EC	0.101%	08:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	08:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:05S		
Control 2 EC	0.100%	08:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:07S		
Control 3 EC	0.100%	08:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:08S	09:08 DJS	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

*Temperature Probe Serial Number: DDWJP2-2042*

**Changed By:**

Last Name: SULLIVAN

First Name: FRANCIS

MI: -

Signature: *Tpr. I J Sullivan #5103*

Badge No.: 5103

Date: 03/30/2009



Drägersafety

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: \_\_\_\_\_

Serial Number:

001553-0210

Certification Date

Technician

Re-Certification Due Date

10-08-2008

JS

10-08-2009

## Drägersafety

### ALCOTEST® 7110 TEMPERATURE PROBE

# CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

Certification date:

Next Certification due:

DDWTP2-204

10-08-2008

10-08-2009

Probe Value

104

Draeger Safety Diagnostics, Inc.  
Technical Service Department

JS



State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 Post Office Box 7068  
 WEST TAUNTON NJ 08628-0068  
 (609) 852-2000

JEN S. CORZINE  
 Governor

ANNE MCGRAN  
 Attorney General

COLONEL JOSEPH R. FLORES  
 Superintendent

**CERTIFICATION OF ANALYSIS  
 0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 03/14/2008

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 08B050

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1229 to 0.1234 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is February 13, 2010.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

*Ajit R. Tungare*  
 Ajit R. Tungare  
 Chief Forensic Scientist  
 Division of State Police

Sworn to and subscribed before me this 31 <sup>st</sup> day of March, 2008.

*Linda L. DeBartis*  
 Linda L. DeBartis  
 My Commission  
 Expires Aug. 17, 2009



New Jersey is An Equal Opportunity Employer  
 Printed on Recycled Paper and Recyclable



DEPARTMENT OF  
**Marine and Public Safety**  
*Marine and Public Safety*

FRANCIS SULLIVAN

New Jersey State Police



IN WITNESS WHEREOF, I have hereunto set my hand and the seal of the State of New Jersey at Trenton, New Jersey, this 17th day of March, 1968.

\_\_\_\_\_  
 FRANCIS SULLIVAN  
 STATE POLICE

\_\_\_\_\_  
 JOHN K. [unclear]  
 STATE POLICE

ORIGINAL COURSE DATES

DATE	INSTRUCTOR	COURSE PLACE	INSTRUCTOR
1. 05/16/68	Essexville, NJ		
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			
10.			

LA. 282 P. 60. 609

DEPARTMENT OF  
**Motor and Public Safety**

Write to be certified that

From Frank J. Sullivan  
NEW JERSEY Police

IN CHALLENGE AND COMPLIANCE TO THE PROVISIONS OF PARAGRAPH 14-B OF THE LAWS OF 1964, AS AMENDED, AND CHAPTER 142 OF THE LAWS OF 1965, AS AMENDED, AND THE 1972 ACTS TO REPEAL AND REENACT THE SAME.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, this 29th day of Apr.  
1974

*Carl Williams*  
SHERIFF

*Robert T. Pitt*  
DEPUTY SHERIFF

ORIGINAL COURSE DATES

No.	DATE	RESTRICTOR COURSE	INSTRUCTOR
1.	9-7-96	ACTC	J. [Signature]
2.	12-15-97	ACTC	J. [Signature]
3.	1-2-98	ACTC	[Signature]
4.	10-9-01	DCPH	[Signature]
5.	7-1-85	ACTC	[Signature]
6.	6-20-87	ACTC	[Signature]
7.	12-16-97	DCPA	[Signature]
8.			
9.			

SP-1038



# CERTIFICATE

This is to certify that


**Francis Sullivan**  
**Trooper I, #5103**

has successfully completed the two day Draeger Safety  
Diagnostics, Inc. Alcohol Coordinator Training Course on  
the New Jersey specific Alcotest® 7110 MKIII-C and is  
hereby certified as a qualified

**Operator Trainer and  
Maintenance Technician**

Completion of this course qualifies this individual to train  
and certify Operators in the proper use and operation as  
well as perform Preventive Maintenance on the New Jersey  
specific Alcotest® 7110 MKIII-C.

Date: September 18, 2008

  
Instructor: Hansueli Byser







Jon S. Corzine  
Governor

State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
PO Box 080  
TRENTON, NJ 08625-0080

ANNE MILGRAM  
Attorney General

November 5, 2008.

Col. Joseph R. Fuentes, Superintendent  
Division of State Police  
Division Headquarters  
P.O. Box 7068  
West Trenton, New Jersey 08628

Re: Breath Test Coordinator/Instructor, Certification - Trooper I Francis Sullivan # 5103

Dear Col. Fuentes:

---

Pursuant to the provisions of N.J.A.C. 13:51-2.1 (b) and (c), as adopted and promulgated under the provisions of N.J.S.A. 39:4-50.3, 39:3-10.25 and 12:7-56, I hereby approve Trooper I Francis Sullivan # 5103 as a duly certified Breath Test Coordinator/Instructor. This approval is effective immediately.

Very truly yours,



Anne Milgram  
Attorney General

AM:HA:tlh

c: Trooper I Francis Sullivan # 5103, Alcohol/Drug Test Unit, Division of State Police  
Lt. Paul Spirit, Unit Head, Alcohol/Drug Test Unit, Division of State Police

Francie Sullivan #5103  
Hamilton Technology Complex  
Suite 400, 1200 Negron Drive  
Hamilton, NJ 08691  
[fp5103@gw.njsp.org](mailto:fp5103@gw.njsp.org)

**Current Rank & Date of Promotion & Time of Service:**

Trooper I                      7/16/2002                      16 years of Service

**Career Experience/Assignment:**

**April 26, 2008 to Present:**

Special Investigation Section  
Alcohol/Drug Testing Unit

**May 29 2004 to April 25 2008:**

Special Operations Section  
Motor Coach/ Compliance Review Unit

**December 29 2001 to May 28 2004:**

Special Operations Section  
Diesel Emissions

**November 3 2001 to December 28 2001:**

Field Operations Section  
Tuckerton Station

Troop A

**April 1 1995 to November 2 2001:**

Field Operations Section  
Bass River Station  
Troop E

**October 15 1994 to March 31 1995:**

Field Operations Section  
Port Norris Station  
Troop A

**October 16 1993 to October 14 1994:**

Field Operations Section  
Tuckerton Station  
Troop A

**January 16, 1993 to October 15 1993:**

Field Operations Section  
Woodbine Station  
Troop A

**Specialized Training:**

Breath Test Coordinator/Instructor, Certification

November 2008

Instructor Training in DWI Detection and Standardized Field Sobriety Testing	October 2008
Draeger Safety Diagnostics, Inc., Alcohol Coordinator Training Course on the New Jersey specific Alcotest 7110 MKIII-C	September 2008
The Robert F. Borkenstein Course on Alcohol and Highway Safety	May 2008
Inspection and Investigation of Commercial Vehicle Crashes	September 2007
Accident Investigation 1	January 2007
Accident Investigation 2	March 2007
Instructor Training Course	May 2006
Laser Operator Training Course	March 2006
Alcotest Certification	March 2006
Post Crash Vehicle Equipment Inspection	November 2004
Compliance Review Training	September 2003
Suspension and Steering	August 2003
Commercial Motor Vehicle Criminal Interdiction	April 2003
General Hazardous Material Inspection	December 2002
Domestic Cannabis Eradication & Suppression Seminar	July 2002
Trucks and Terrorism	May 2002
Motorcoach	May 2002
NASI-North American Standard Inspection Part B	February 2002
NASI-Weights & Dimensions	February 2002
NASI-North American Standard Inspection Part A	January 2002
PR-24 Training Course	July 1999
Narcotics Trafficking Trends Seminar	June 1998
Firearms Instructor Course	May 1997
Driving While Intoxicated/Standardized Field Sobriety	July 1996

**Educational/Certification Summary:**

Associates of Science	
Comm. College of the Air Force	August 1987

**Commendation & Awards:**  
Certificate of Commendation

June 2002

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
<b>Location:</b>	BUENA BOROUGH POLICE		
Calibration File No.:	00004	Calib. Date:	11/05/2008
Certification File No.:	00005	Cert. Date:	11/05/2008
Linearity File No.:	00006	Lin. Date:	11/05/2008
Solution File No.:	00019	Soln. Date:	03/07/2009
Sequential File No.:	00019	File Date:	03/07/2009
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXJ S3-0209
Solution Control Lot:	08B050	Expires:	02/13/2010
		Bottle No.:	0481

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:48S		
Control 1 EC	0.102%	18:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:49S		
Control 2 EC	0.100%	18:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:50S		
Control 3 EC	0.100%	18:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature: 

Badge No.: 1106

Date: 03/07/2009

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C		Serial No.: ARXB-0074
Location: BUENA BOROUGH POLICE		
Calibration File No.: 00004	Calib. Date: 11/05/2008	Calib. No.: 00002
Certification File No.: 00005	Cert. Date: 11/05/2008	Cert. No.: 00001
Linearity File No.: 00006	Lin. Date: 11/05/2008	Lin. No.: 00001
Solution File No.: 00018	Soln. Date: 02/03/2009	Soln. No.: 00007
Sequential File No.: 00018	File Date: 02/03/2009	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXJ S3-0209
Control Solution %: 0.100%		Expires: 02/13/2010
Solution Control Lot: 08B050		Bottle No.: 0486

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:24S		
Control 1 EC	0.103%	20:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	20:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:25S		
Control 2 EC	0.101%	20:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	20:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:27S		
Control 3 EC	0.101%	20:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature: \_\_\_\_\_



Badge No.: 1106

Date: 02/03/2009

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
Location:	BUENA BOROUGH POLICE	
Calibration File No.:	00004	Calib. Date: 11/05/2008
Certification File No.:	00005	Cert. No.: 00001
Linearity File No.:	00006	Lin. Date: 11/05/2008
Solution File No.:	00016	Soln. No.: 00006
Sequential File No.:	00016	File Date: 12/31/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXJ S3-0209
Solution Control Lot:	08B050	Expires: 02/13/2010
		Bottle No.: 0491

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:58S		
Control 1 EC	0.102%	19:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:00S		
Control 2 EC	0.100%	20:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	20:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:01S		
Control 3 EC	0.100%	20:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature: \_\_\_\_\_



Badge No.: 1106

Date: 12/31/2008

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
Location:	BUENA BOROUGH POLICE	
Calibration File No.:	00004	Calib. Date: 11/05/2008
Certification File No.:	00005	Cert. No.: 00002
Linearity File No.:	00006	Cert. Date: 11/05/2008
Solution File No.:	00013	Lin. No.: 00001
Sequential File No.:	00013	Soln. Date: 11/30/2008
		Soln. No.: 00005
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXJ S3-0209
Solution Control Lot:	08B050	Expires: 02/13/2010
		Bottle No.: 0488

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:53S		
Control 1 EC	0.100%	09:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54S		
Control 2 EC	0.099%	09:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56S		
Control 3 EC	0.100%	09:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRILLO

First Name: LAWRENCE

MI: A

Signature: 

Badge No.: 1106

Date: 11/30/2008