

Stratford Police Department

Portable Breath Test Unit Accuracy Verification / Calibration Log

Unit number	Date	Verification Results	Calibration (Yes/No)	Post calibration Results	Notes	Operator Name
SPD 1	01/17/16	.071%	YES	.079%	batteries changed, functioning properly, label replaced, returned to service	Chief R. Morello
SPD 2	01/17/16	.079%	NO	NO CAL	batteries changed, functioning properly label replaced returned to service	Chief R. Morello
SPD 3	01/17/16	.075%	NO	NO CAL	batteries changed, functioning properly, returned to service	Chief R. Morello
SPD 4	01/17/16	.076%	NO	NO CAL	batteries changed, functioning properly, returned to service	Chief R. Morello
SPD 5	01/17/16	.072%	YES	.076%	batteries changed, functioning properly, returned to service	Chief R. Morello
Sgt Kelly	01/17/16	.068%	YES	.078%	batteries changed, functioning properly	Chief R. Morello
Sgt Gagliardi	01/17/16	.079%	NO	NO CAL	batteries changed, functioning properly	Chief R. Morello
SPARE 3	01/17/16	.071%	NO	NO CAL	batteries changed, functioning properly	Chief R. Morello
Schmidt	01/17/16	.076%	NO	NO CAL	batteries changed, functioning properly, returned to service	Chief R. Morello