

Stratford Police Department

Portable Breath Test Unit Accuracy Verification / Calibration Log

Unit number	Date	Verification Results	Calibration (Yes/No)	Post calibration Results	Notes	Operator Name
SPD 1	6/21/15	.060%	YES	.079%	batteries changed, functioning properly, returned to service	Chief R. Morello
SPD 2	6/21/15	.072%	YES	.076%	batteries changed, functioning properly, returned to service	Chief R. Morello
SPD 3	6/21/15	.075%	NO	NO CAL.	batteries changed, functioning properly, returned to service	Chief R. Morello
SPD 4	6/21/15	.075%	NO	NO CAL	batteries changed, functioning properly, returned to service	Chief R. Morello
SPD 5	6/21/15	.066%	YES	.077%	batteries changed, functioning properly, returned to service	Chief R. Morello
SPARE 1	6/21/15	.065%	YES	.080%	batteries changed, functioning properly	Chief R. Morello
SPARE 2	6/21/15	.07%	YES	.080%	batteries changed, functioning properly	Chief R. Morello
SPARE 3	6/21/15	.068%	YES	.080%	batteries changed, functioning properly	Chief R. Morello
Schmidt	6/21/15	.066%	YES	.079%	batteries changed, functioning properly, returned to service	Chief R. Morello