

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Eastampton Township Police Department 725 Smithville Road Eastampton, NJ 08060		2. Contact: Dennis Shepard	4. Date: 02-12-2015
		3. Phone Number: 609-261-1717	5. County: Burlington
6. Alcotest Instrument Serial Number: ARNK-0038	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
During a linearity test the screen displayed "ERROR (035) EC-SYSTEM".

10. Comments:
Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE


Last Known Sequential File #: 00482

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Sgt. Michelle L. Goncalves #6040
 Name & Badge Number (Print)

2/12/15


 Signature

Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Eastampton Twp. P.D., Mount Holly, NJ

Date Received: 04/13/15

Date given to service: 04/13/15

Carrier: FedEx UPS USPS

Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR NK-0038

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: October 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARFN-0053

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 04/23/15

Packing Slip

Customer no Order No Order date
 150053294 10773629 04/23/2015

Packing slip no. Ship date
 80941895 04/23/2015

Please reference on inquiries

Ship to

150053296
 EASTAMPTON TOWNSHIP POLICE DEPT.
 ATTN: LT. TOM DURHAM
 725 SMITHVILLE ROAD
 MOUNT HOLLY, NJ 08060
 USA

Payer

Your Purchase Order Number

DENNIS SHEPARD
 Any questions? Please contact:
 ROSEMARIE HUGHES

Customer
 EASTAMPTON TOWNSHIP
 POLICE DEPARTMENT
 BURLINGTON COUNTY
 725 SMITHVILLE ROAD
 MOUNT HOLLY, NJ 08060
 USA

Your sales person

LINDA SALO
 Phone: 866-385-5900
 Fax: 972-929-1260

Delivery terms
 FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARNK-0038 REC'D 4/13/15			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA			
0020	0.5 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 EA			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 536410
 Pittsburgh, PA 15253-5906

Packing Slip

Customer no Order No Order date
150053294 10773629 04/23/2015

Packing slip no. Ship date
80941895 04/23/2015
Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			
0040	2 EA	6808486 PLATES,7110--9510 Ordered / Back ordered : 0 / 0 EA			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Eastampton Township Police Department 725 Smithville Road Eastampton, NJ 08060	2. Contact: Dennis Shepard	4. Date: 05-07-2015
	3. Phone Number: 609-261-1717	5. County: Burlington

6. Alcotest Instrument Serial Number: ARNK-0038	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A
---	---	---

9. Reason for Service:
The above Alcotest was returned from outside evaluation and/or repair. A successful calibration was conducted utilizing the above Alcotest.

10. Comments:
See attached packing slip and Drager Return and Repair Form.

STATE OF NEW JERSEY

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Sgt. Michelle L. Goncalves #6040
Name & Badge Number (Print)

Sgt. Michelle L. Goncalves #6040 5/7/15
Signature Date

Packing Slip

Customer no Order No Order date
 150053294 10271619 03/12/2008

Packing slip no. Ship date
 80354203 03/12/2008

Please reference on inquiries

ship to

150053296
 EASTAMPTON TOWNSHIP POLICE DEPT.
 ATTN: LT. TOM DURHAM
 725 SMITHVILLE ROAD
 MOUNT HOLLY, NJ 08060
 USA

bill to

Your Purchase Order Number

R. COPPOLA
 Any questions? Please contact:
 MARA HERRERA

Customer
 EASTAMPTON TOWNSHIP
 POLICE DEPARTMENT
 BURLINGTON COUNTY
 725 SMITHVILLE ROAD
 MOUNT HOLLY, NJ 08060
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Econ (2day)			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110 SER# ARNK-0038			
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			
0020	1 EA	MPCAL71 CALIBRATION CHARGE - 7110 Ordered / Back ordered : 0 / 0 EA			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150053294 10271619 03/12/2008

Packing slip no. Ship date
80354203 03/12/2008

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

Dräger safety

Internal Use Only

Return & Repair Form

Customer Information:

CUSTOMER TOWNSHIP OF EASTAMPTON P. D. Cust # _____
 Date of Return: 03/07/2008 Carrier: UPS DHL USPS Method: GRD NDA ZDY 3DY
 OTHER: _____

Product Information:

Product: 7110 7410 6510 6810 Serial # AR NM - 0038
 Description: A B PLUS SCREENER Printer Serial # AR - _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # DD P - _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: _____

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL 71</u>	<u>7110 CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPC LABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes: REPLACE FC / CAL w/ QC & OPS CHECKS

RAN 50 ADDITIONAL TESTS LOOKING FOR IR SPIKES -
NONE NOTED

Technician: [Signature]

Date: 03-11-08