

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Eastampton PD 1380 Woodlane Road, Eastampton NJ 08060		2. Contact: Lt. Dennis Shepard	4. Date: 1/31/2020
		3. Phone Number: 609-261-1717	5. County: Burlington
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARNK-0038</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>	

9. Reason for Service:
Complaint received of the inability to remotely connect to the Alcotest instrument on (3) three separate occasions while attempting to retrieve data.

10. Comments:
The instrument was placed out of service and returned to Draeger Inc. for evaluation.

STATE OF NEW JERSEY

Last Known Sequential File #: 630

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dave Napolitano #7237
Name & Badge Number (Print)

[Signature] 237 1-31-20
Signature Date



Delivery receipt



Customer no.
91266631

Number of report
304922000
Please reference on Inquiries

Date of receiving Delivery receipt date
/ / 04/10/2020

Consignee, 91267771
EASTHAMPTON POLICE DEPT
PO#REQ.R2000043 / ARNK-0038
1380 WOODLANE RD
MOUNT HOLLY NJ 08060-9696

Customer, 91266631
EASTAMPTON POLICE DEPT
PO#REQ.R2000043 / ARNK-0038
1380 WOODLANE RD
MOUNT HOLLY NJ 08060-4379

Your order

Date of order:
02/06/2020
Purch.ord.no.:
20-00128

Branch text

Dräger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
	ALCOTEST 7110 MK III USA	8314246	1030033773	ARNK-0038
	REPAIR/ARNK-0038			

Dräger Inc.
Our Tax ID: 23-1699096
3135 Quamy Road, Telford, PA 18969
An Equal Opportunity Employer M/F/V/H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Dräger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Dräger Inc.
Account Number: 00-494-836
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no. 91266631
 Number of report 304922000
 Date of receiving
 Please reference on inquiries

Customer
 EASTAMPTON POLICE DEPT
 12 MANOR HOUSE CT
 MOUNT HOLLY NJ 08060-4379

Consignee
 EASTHAMPTON POLICE DEPT
 1380 WOODLANE RD
 MOUNT HOLLY NJ 08060-9696

Your order

Date of order:
 02/06/2020
 Your reference:
 20-00128

NL-Text

Your contact person

Service Support-IDT
 800-437-2437 OPT 5
 SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 142538870
		REPAIR/ARNK-0038
		Eq.No. : I030033773 SN: ARNK-0038
		Material: 8314246 ALCOTEST 7110 MK III USA
		Inv.No. :
		Location:
		REPAIR/ARNK-0038
1.5	H	R001 Repair time - normal hours
1	EA	6808374 Geared Engine, A7110/A9510
1	EA	6808455 Sensor Alcotest B

Ref:	Date: 13Apr20	SHIPPING:	32.71
Dep:	Wgt: 18.00 LBS	SPECIAL:	0.85
	DV:	HANDLING:	0.00
	Svc: STANDARD OVERNIGHT	TOTAL:	33.36
	TRCK-1395 7032 2436		

Draeger Inc.
 Our Tax ID: 23-1699096
 3135 Quarry Road; Telford, PA 18969

Remit to:
 LOCKBOX (Standard USPS)
 Draeger, Inc.

Remit to:
 LOCKBOX (Overnight)
 FIS Lockbox Processing

Remit US Wire Transfers to:
 Account Name: Draeger Inc.
 Account Number: 00-494-936

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Eastampton PD 1380 Woodlane Road, Eastampton NJ 08060		2. Contact: Lt. Dennis Shepard	4. Date: 1/31/2020
		3. Phone Number: 609-261-1717	5. County: Burlington
6. Alcotest Instrument Serial Number: ARNK-0038	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
Complaint received of the inability to remotely connect to the Alcotest instrument on (3) three separate occasions while attempting to retrieve data.

10. Comments:
The instrument was placed out of service and returned to Draeger Inc. for evaluation.

STATE OF NEW JERSEY
1921

Last Known Sequential File #: 630

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dave Napolitano #7237
Name & Badge Number (Print)

[Signature]
Signature

1-31-20
Date

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Eastampton PD 1380 Woodlane Road, Eastampton NJ 08060	2. Contact: Lt. Dennis Shepard	4. Date: 4/29/2020
	3. Phone Number: 609-261-1717	5. County: Burlington
6. Alcotest Instrument Serial Number: ARNK-0038	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
The alcotest instrument was returned from outside evaluation and placed back in service.

10. Comments:
See Draeger Return and Repair form.


STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
1922

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I David M. Napolitano #7237
Name & Badge Number (Print)


Signature

4-29-20
Date

RETURN AND REPAIR FORM



Customer Information

B: _____

S: _____

Company Name: Eastampton, NJ

Date Received: 02/24/2020

Date given to service: 02/24/2020

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR NK-0038

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	
MP Labor	Labor	1.5	
6808455	Fuel Cell	1	
6808374	Geared Engine	1	

Repair Notes: Replaced defective fuel cell, geared engine, and repaired modem board.

FUEL CELL = ARMM-1283

CAL W/QC AND OPS CHECK

Service Technician: _____

Date: 04/09/2020

Service note

Customer no. 91266631 Number of report 304922000 Date of receiving
 Please reference on inquiries

Customer
 EASTAMPTON POLICE DEPT
 12 MANOR HOUSE CT
 MOUNT HOLLY NJ 08060-4379

Consignee
 EASTHAMPTON POLICE DEPT
 1380 WOODLANE RD
 MOUNT HOLLY NJ 08060-9696

Your order

Date of order:
 02/06/2020
 Your reference:
 20-00128

NL-Text

Your contact person

Service Support-IDT
 800-437-2437 OPT 5
 SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 142538870
		REPAIR/ARNK-0038
		Eq.No. : 1030033773 SN: ARNK-0038
		Material: 8314246 ALCOTEST 7110 MK III USA
		Inv.No. :
		Location:
		REPAIR/ARNK-0038
	1.5 H	R001 Repair time - normal hours
	1 EA	6808374 Geared Engine, A7110/A9510
	1 EA	6808455 Sensor Alcotest B



Delivery receipt



Customer no.
91266631

Number of report
304922000

Date of receiving Delivery receipt date
/ / 04/10/2020

Please reference on inquiries

Consignee, 91267771
EASTHAMPTON POLICE DEPT
PO#REQ.R2000043 / ARNK-0038
1380 WOODLANE RD
MOUNT HOLLY NJ 08060-9696

Customer, 91266631
EASTAMPTON POLICE DEPT
PO#REQ.R2000043 / ARNK-0038
1380 WOODLANE RD
MOUNT HOLLY NJ 08060-4379

Your order

Date of order:
02/06/2020
Purch.ord.no. :
20-00128

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
	ALCOTEST 7110 MK III USA	8314246	1030033773	ARNK-0038
	REPAIR/ARNK-0038			

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road; Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

COE005738

Dräger

Acknowledgement



Customer no.
91266631

Number of report
304747065

Date of receiving
/ /

Please reference on inquiries

Customer

EASTAMPTON POLICE DEPT
ATTN CAROL STONE
PO# DDUKS3-0131
1380 MANOR HOUSE CT
MOUNT HOLLY NJ 08060-4379
USA

Consignee

EASTHAMPTON POLICE DEPT
ATTN CAROL STONE
PO# DDUKS3-0131
1380 WOODLANE RD
MOUNT HOLLY NJ 08060-9696
USA

Your order

Date of order

Purch. ord. no.

Branch text

Your contact person

STACEY COOK
800-437-2347 OPT 5
CS-IDT@draeger.com

Page 1

Item	Quantity	Description	Part-No.	Equipment-No.	Serial-No.
10	001	Simulator, CU34 110V elec ctr	4540083	1020891259	DDUKS3-0131
20	001	Simulator, temperature probe	4412025	1020891260	DDUNP2-253

Date of delivery (CW/YY): 47/19

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road, Telford, PA 18969
An Equal Opportunity Employer M/F/V/H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005