

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Burlington Twp. P.D. 551 Old York Rd Burlington, NJ 08616	2. Contact: Sgt Shopper	4. Date: 3/15/10
	3. Phone Number: 609-386-2019	5. County: Burlington
6. Alcotest Instrument Serial Number: ARTL-0024	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
Alcotest returned from outside repair and placed back in service.

10. Comments:
Fuel cell replaced.
See Draeger packing slip.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr II Steve Dellenice 6027 Tpr II [Signature] 6027 3/15/10
Name & Badge Number (Print) Signature Date

Packing Slip

Customer no Order No Order date
 150053898 10394530 03/05/2010

Packing slip no. Ship date
 80503550 03/05/2010

Please reference on inquiries

ship to

150053898
 BURLINGTON TOWNSHIP POLICE DEPT.
 ATTN: SGT. KEVIN SHOPPAS
 851 OLD YORK ROAD
 BURLINGTON, NJ 08016
 USA

bill to

Your Purchase Order Number

7110 WARRANTY: ARTL-0024
 Any questions? Please contact:
 MARA HERRERA

Customer

BURLINGTON TOWNSHIP POLICE DEPT.
 ATTN: SGT. KEVIN SHOPPAS
 851 OLD YORK ROAD
 BURLINGTON, NJ 08016
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms

FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARTL-0024			
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA REPLACED			
0020	1 EA	MPCAL71 CALIBRATION CHARGE - 7110 Ordered / Back ordered : 0 / 0 EA			

Dräger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209428615
 Acct. Name: Dräger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Dräger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150053898 10394530 03/05/2010

Packing slip no. Slip date
80503550 03/05/2010

Please reference on Inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 FT	8312074 TUBING, INTAKE FOR 7410 Ordered / Back ordered : 0 / 0 FT REPLACED			
0040	1 EA	6808486 PLATES, 7110 Ordered / Back ordered : 0 / 0 EA REPLACED			
0050	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

Dräger safety

Alcotest[®] 7110 MKIII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864 and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:

03/04/2000

SERIAL NUMBER:

APR 16-0084

Dräger Safety Diagnostics, Inc.



STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: <i>Burlington Twp Police Dept 851 Old York Rd Burlington, NJ 08016</i>	2. Contact: <i>Sgt Shepper</i>	4. Date: <i>2/24/10</i>
	3. Phone Number: <i>609-396-2019</i>	5. County: <i>Burlington</i>

6. Alcotest Instrument Serial Number: <i>ARTL-0024</i>	7. Simulator Component Serial Number: <i>—</i>	8. Temperature Probe Component Serial Number: <i>—</i>
---	---	---

9. Reason for Service:
At scheduled semi-annual calibration fuel cell replacement was deemed to be required.

10. Comments:
Please replace fuel cell.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II JS Dellenore 6027
Name & Badge Number (Print)

Tpr II JS # 6027
Signature *2/24/10*
Date