

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Longport Police Department 2305 Atlantic Avenue Longport, NJ 08403		2. Contact: Capt Vince Pacentrilli	4. Date: 08/13/2010
		3. Phone Number: 609-822-2141	5. County: Atlantic
6. Alcotest Instrument Serial Number: ARXA-0067	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
Alcotest instrument returned from outside maintenance and placed back in service.

10. Comments:
Fuel cell replaced. See Draeger packing slip.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Steven Dellanoce #6027
 Name & Badge Number (Print)

Tpr II Steven Dellanoce #6027
 Signature

8/13/10
 Date

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Longport Police Department 2305 Atlantic Avenue Longport, NJ 08403		2. Contact: Capt Vince Pacentrilli	4. Date: 08/03/2010
		3. Phone Number: 609-822-2141	5. County: Atlantic
6. Alcotest Instrument Serial Number: ARXA-0067	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
While reviewing the performance history of the instrument prior to recalibration low EC results on control tests were noted.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed low Electrochemical Sensor (EC) results (Fuel Cell) on control tests. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Please replace Fuel Cell.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Steven Dellanoce #6027
Name & Badge Number (Print)

Tpr. II Steven Dellanoce #6027
Signature

8/3/10
Date

Packing Slip

Customer no. Order No Order date
 150062710 10423822 08/10/2010

Packing slip no. Ship date
 80538795 08/10/2010

Please reference on inquiries

ship to

150062710
 BOROUGH OF LONGPORT POLICE
 ATTN: CAPTAIN PACENTRILLI
 2305 ATLANTIC AVENUE
 ATLANTIC COUNTY
 LONGPORT, NJ 08403
 USA

bill to

Your Purchase Order Number

7110 WARRANTY: ARXA-0067
 Any questions? Please contact:
 MARA HERRERA

Customer
 BOROUGH OF LONGPORT POLICE
 ATTN: CAPTAIN PACENTRILLI
 2305 ATLANTIC AVENUE
 ATLANTIC COUNTY
 LONGPORT, NJ 08403
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms

FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110: ARXA-0067
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA REPLACED
0020	1 EA	6808486 PLATES, 7110 Ordered / Back ordered : 0 / 0 EA REPLACED

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337