

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Florence Township Police Department 711 Broad Street Florence, NJ 08518		2. Contact: Captain B. Boldizar	4. Date: 12-29-2016
		3. Phone Number: 609-449-0700	5. County: Burlington
6. Alcotest Instrument Serial Number: ARTN-0053	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
 Two recalibration attempts resulted in control test failures for EC readings not being within tolerance.

10. Comments:
 Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #: 0893

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Sgt. Michelle L. Goncalves #6040
 Name & Badge Number (Print)

Sgt. Michelle L. Goncalves #6040 #6040 12/29/16
 Signature Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Florence Twp. P.D., Florence, NJ

Date Received: 01/09/2017

Date given to service: 01/09/2017

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR TN-0053

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARJE-1730

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 01/10/2017

Packing Slip

Customer no Order No Order date
150054650 10858157 01/10/2017

Packing slip no. Ship date
81035871 01/10/2017
Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			
0040	2 EA	6808486 PLATES, 7110--9510 Ordered / Back ordered : 0 / 0 EA			

Packing Slip

Customer no Order No Order date
 150054650 10858157 01/10/2017

Packing slip no. Ship date
 81035871 01/10/2017
 Please reference on inquiries

Ship to
 150054650
 TOWNSHIP OF FLORENCE
 ATTN: DSGT. AL JACOBY
 BURLINGTON COUNTY
 711 BROAD STREET
 FLORENCE, NJ 08518-2310
 USA

Payer

Your Purchase Order Number

DSGT. AL JACOBY
 Any questions? Please contact:
 ROSEMARIE HUGHES

Customer
 TOWNSHIP OF FLORENCE
 BURLINGTON COUNTY
 711 BROAD STREET
 FLORENCE, NJ 08518-2310
 USA

Your sales person

CHRISTOPHER FAIRCHILD
 Phone: 866-385-5900
 Fax: 972-929-1260

Delivery terms
 FA
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110:ARTN-0053 REC'D 1/8/17
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA
0020	0.5 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 EA

Draeger Safety Diagnostics Inc.
 Accounting Address:
 3135 Quarry Road
 Telford, PA 18969 USA
 Toll-free: +1-800-437-2437
 Fax: +1-215-721-5410

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 536410
 Pittsburgh, PA 15253-5908

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Florence Township Police Department 711 Broad Street Florence, NJ 08518		2. Contact: Captain B. Boldizar	4. Date: 01-20-2017
		3. Phone Number: 609-449-0700	5. County: Burlington
6. Alcotest Instrument Serial Number: ARTN-0053	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
The above Alcotest was returned from outside evaluation and/or repair. A successful calibration was conducted utilizing the above Alcotest.


10. Comments:
See attached packing slip and Drager Return and Repair Form.

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Sgt. Michelle L. Goncalves #6040
Name & Badge Number (Print)


Signature

#6040 1/29/17
Date

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Florence Twp Police Department 711 Broad St Florence, NJ 08518		2. Contact: Capt Scully	4. Date: 03/10/11
		3. Phone Number: 609-499-3131	5. County: Burlington
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARTN-0053</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>	

9. Reason for Service:
Complaint received of instrument experiencing repeated control test failures.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed several repeated control test failures due to low Electrochemical Sensor (EC) results. Instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Please replace FUEL CELL.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Steve Dellanoce #6027 <small>Name & Badge Number (Print)</small>	<small>Signature</small>		#6027 3/10/11 <small>Date</small>
--	--------------------------	--	---

Packing Slip

Customer no Order No Order date
 150054650 10465022 03/21/2011

Packing slip no. Ship date
 80587585 03/21/2011
 Please reference on inquiries

ship to
 150054650
 TOWNSHIP OF FLORENCE
 BURLINGTON COUNTY
 711 BROAD STREET
 FLORENCE, NJ 08518
 USA

bill to

Your Purchase Order Number
 03/21/2011
 ARTN-0053 WARRANTY
 Any questions? Please contact:
 ERIK ESQUEDA

Customer
 TOWNSHIP OF FLORENCE
 BURLINGTON COUNTY
 711 BROAD STREET
 FLORENCE, NJ 08518
 USA

sales person
 Don Poulict
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Ground			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110 SERIAL #: ARTN-0053 FUEL CELL WAS REPLACED. FUEL CELL #: ARBM-3279. 7110 EXTENDED WARRANTY EXPIRES: OCTOBER 2013.			
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 1 / 0 EA **REPLACED**			

Dräger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209428815
 Acct. Name: Dräger Safety Diagnostics Inc.
 Transit Number: 038078150
 Federal ID Number: 84-1600159

REMIT TO:
 Dräger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150054650 10465022 03/21/2011

Packing slip no. Ship date
80587585 03/21/2011

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0020	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 1 / 0 EA			
0030	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.5 / 0.0 H			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Florence Twp Police Department 711 Broad St Florence, NJ 08518		2. Contact: Capt Scully	4. Date: 03/28/11
		3. Phone Number: 609-499-3131	5. County: Burlington
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARTN-0053</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>	

9. Reason for Service:
Alcotest instrument returned from outside maintenance and placed back in service.

10. Comments:
Fuel cell replaced. See Draeger packing slip.

- 11. The above instrument/component has been found to be in satisfactory working condition; no further action required.
- 12. The above instrument/component is placed out of service pending further evaluation.
- 13. The above instrument/component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Steve Dellanoce #6027 <small>Name & Badge Number (Print)</small>	 <small>Signature</small>	3/28/11 <small>Date</small>
--	------------------------------	---------------------------------------

Dräger

Alcotest® 7110 MKIII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864 and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:

SERIAL NUMBER:

3/21/2011

ARTN-0053

Dräger Safety Diagnostics, Inc.



STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Florence Police Department 711 Broad Street Florence NJ, 08518		2. Contact: Chief Brian Boldizar	4. Date: 2/21/2020
		3. Phone Number: 609-499-3131	5. County: Burlington
6. Alcotest Instrument Serial Number: ARTN-0053	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
Complaint received of the instrument experiencing multiple control test failures.


10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed control test failures due to low electrochemical sensor (EC) results. The instrument was placed out of service and returned to Draeger Inc. for evaluation.

Last Known Sequential File #: 1080

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I David M. Napolitano #7237
Name & Badge Number (Print)


Signature

2-21-20
Date

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Florence Twp PD 711 Broad Street Florence, NJ 08518	2. Contact: Chief Brian Boldizar	4. Date: 4/6/2020
3. Phone Number: 609-499-0700		5. County: Burlington
6. Alcotest Instrument Serial Number: ARTN-0053	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
The alcotest instrument was returned from outside evaluation and placed back in service.


10. Comments:
See Draeger Return and Repair form.

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I David M. Napolitano #7237
Name & Badge Number (Print)


Signature

4.6.20
Date

RETURN AND REPAIR FORM



Customer Information

B: _____

S: _____

Company Name: Florence Police Department, Florence, NJ

Date Received: 3/12/2020

Date given to service: 3/12/2020

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR TN-0053

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information

Test#: _____

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	
MP Labor	Labor	1	
6808455	Fuel Cell	1	
6808374	Geared Engine	1	

Repair Notes: Replaced defective fuel cell and geared engine.

FUEL CELL = ARMM-1240
CAL W/QC AND OPS CHECK

Service Technician:

Date: 03/25/2020

Delivery receipt



Customer no.
150054650

Number of report
304992493

Date of receiving Delivery receipt date
/ / 03/26/2020

Please reference on inquiries

Consignee, 150054650
TOWNSHIP OF FLORENCE
MUNICIPAL COMPLEX
PO#20-00474
ATTN: LT. JACOBY
711 BROAD ST
FLORENCE NJ 08518-2310

Customer, 150054650
TOWNSHIP OF FLORENCE
MUNICIPAL COMPLEX
PO#20-00474
ATTN: LT. JACOBY
711 BROAD ST
FLORENCE NJ 08518-2310

Your order

Date of order:
03/05/2020
Purch.ord.no.:
20-00474

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
	ALCOTEST 7110 MK III USA	8314246	1030356399	ARTN-0053
	REPAIR/ARTN-0053			

Service note

Customer no. 150054650 Number of report 304992493 Date of receiving

Please reference on inquiries

Customer
TOWNSHIP OF FLORENCE
MUNICIPAL COMPLEX
711 BROAD ST
FLORENCE NJ 08518-2310

Consignee
TOWNSHIP OF FLORENCE
MUNICIPAL COMPLEX
711 BROAD ST
FLORENCE NJ 08518-2310

Your order

Date of order:

03/05/2020

Your reference:

20-00474

NL-Text

Your contact person

Service Support-IDT

800-437-2437 OPT 5

SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 142635565
		REPAIR/ARTN-0053
		Eq.No. : 1030356399 SN: ARTN-0053
		Material: 8314246 ALCOTEST 7110 MK III USA
		Inv.No. :
		Location:
		REPAIR/ARTN-0053
	1.0 H	R001 Repair time - normal hours
	1 EA	6808455 Sensor Alcotest B
	1 EA	6808374 Geared Engine, A7110/A9510