

STATE OF NEW JERSEY  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Buena Borough Police Department          616 Central Ave.          Minotola, NJ 08341</b>	2. Contact: <b>Sgt. Petrilla</b> 3. Phone Number: <b>856-697-2810</b>	4. Date: <b>8/31/10</b> 5. County: <b>Atlantic</b>
------------------------------------------------------------------------------------------------------------------------	--------------------------------------------------------------------------------	-------------------------------------------------------------

6. Alcotest Instrument Serial Number: <b>ARXB-0074</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>
-----------------------------------------------------------	-----------------------------------------------------	-------------------------------------------------------------

9. Reason for Service:  
**Alcotest instrument returned from outside maintenance and placed back in service.**


10. Comments:  
**Fuel cell replaced. Cuvette replaced. Barb, brass, metric w/ valve replaced.  
 See Draeger packing slip.**

**Note: The Alcotest data from 03/30/10 until 08/13/10 was not retrieved by Draeger.**

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Steve Dellanoc #6027  
 Name & Badge Number (Print)
 

  
 Signature
 

8/31/10  
 Date

# Packing Slip

Customer no    Order No    Order date  
 150065405    10425715    08/19/2010

Packing slip no.    Ship date  
 80540965    08/19/2010

Please reference on inquiries

ship to  
 150065406  
 BOROUGH OF BUENA POLICE DEPARTMENT  
 ATTN: CHIEF ADAMS  
 616 CENTRAL AVENUE  
 ATLANTIC COUNTY  
 MINOTOLA, NJ 08341  
 USA

bill to

Your Purchase Order Number

7110 WARRANTY; ARXB-0074  
 Any questions? Please contact:  
 MARA HERRERA

Customer  
 BOROUGH OF BUENA  
 ATTN: FINANCE DEPARTMENT  
 P.O. BOX 620  
 ATLANTIC COUNTY  
 MINOTOLA, NJ 08341-0620  
 USA

Your sales person  
 Don Pouliot  
 Phone: 973-398-3228

Delivery terms  
 FA  
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARXB-0074			
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA REPLACED			
0020	1 EA	6809500 CUVETTE, ALCOTEST PV LARGE PREWARM TRAC Ordered / Back ordered : 0 / 0 EA REPLACED			

**Draeger Safety Diagnostics Inc.**  
 Accounting Address:  
 101 Technology Drive  
 Pittsburgh, PA 15275  
 Tel: (412) 788-5537  
 Fax: (412) 788-5588

**Remit Wire Transfers To:**  
 Citizens Bank  
 Acct. # 6209426615  
 Acct. Name: Draeger Safety Diagnostics Inc.  
 Transit Number: 036076150  
 Federal ID Number: 84-1600159

**REMIT TO:**  
 Draeger Safety Diagnostics Inc.  
 P.O. Box 200337  
 Pittsburgh, PA 15251-0337

# Packing Slip

Customer no    Order No    Order date  
150065405    10425715    08/19/2010

Packing slip no.    Ship date  
80540965    08/19/2010

Please reference on Inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6809512 BARB, BRASS, METRIC W/ VALVE Ordered / Back ordered : 0 / 0 EA REPLACED			
0040	1 EA	MPCAL71 CALIBRATION CHARGE - 7110 Ordered / Back ordered : 0 / 0 EA			
0050	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Buena Borough Police Department 616 Central Ave. Minotola, NJ 08341</b>		2. Contact: <b>Sgt. Petrilla</b>		4. Date: <b>08/13/2010</b>	
		3. Phone Number: <b>856-697-2810</b>		5. County: <b>Atlantic</b>	
6. Alcotest Instrument Serial Number: <p style="text-align: center;"><b>ARXB-0074</b></p>		7. Simulator Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>		8. Temperature Probe Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>	

9. Reason for Service:  
**Complaint received from Buena Borough Police of ERROR [023] IR-SYSTEM on display screen of Alcotest.**

10. Comments:  
**Operational condition of the instrument was assessed. Upon arrival at the police department the instrument displayed ERROR [023] IR-SYSTEM. I powered the instrument off and back on and heard an audible series of beeps then ERROR [071] HEATER SYSTEM appeared on the display screen. I was not able to clear this message from display screen.**

**I then observed the simulator and temperature probe. The simulator jar was extremely hot to the touch, the heating element glowed bright red, and approximately one-quarter of the jar still contained solution which was murky due to the melted plastic baffle. I was unable to pull any reports from the memory but was provided with a solution change report from 08/09/10 which revealed a failure due to a simulator temp error of 72.4 degrees C.**

**NOTE: I was unable to download the data due to the display screen being locked on ERROR [071] HEATER SYSTEM.**

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<p><b>Tpr. II Steve Dellanoce #6027</b>          _____          Name &amp; Badge Number (Print)</p>	<p><i>Tpr II Steve Dellanoce #6027</i>          _____          Signature</p>	<p><b>8/13/10</b>          _____          Date</p>
-------------------------------------------------------------------------------------------------------------	--------------------------------------------------------------------------------------	------------------------------------------------------------

BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: <i>Bureau Boro Police Dept</i>		2. Contact: <i>Sgt Petrillo</i>		4. Date: <i>9/24/10</i>
3. Phone Number: <i>609-697-2810</i>		5. County: <i>Atlantic</i>		

6. Alcotest Instrument Serial Number: <i>N/A</i>	7. Simulator Component Serial Number: <i>DDRT-0006</i>	8. Temperature Probe Component Serial Number: <i>DDWTFa-088</i>
--------------------------------------------------	--------------------------------------------------------	-----------------------------------------------------------------

9. Reason for Service: *Simulator and temperature probe components returned from outside maintenance and placed back in service.*

10. Comments: *Components rechecked. See Drawer packing slip.*

- 11. The above instrument/component has been found to be in satisfactory working condition; no further action required.
- 12. The above instrument/component is placed out of service pending further evaluation.
- 13. The above instrument/component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the foregoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the foregoing statements made by me are willfully false, I am subject to punishment."

Name & Badge Number (Print): <i>Tr II Steve DeLance # 6027</i>	Signature: <i>Tr II Steve DeLance # 6027</i>
	Date: <i>9/24/10</i>

# Packing Slip

Customer no 150065405 Order No 10429212  
 Order date 09/08/2010

Packing slip no 80544982  
 Ship date 09/08/2010  
 Please reference on inquiries

Ship to  
 150065406  
 BOROUGH OF BUENA POLICE DEPARTMENT  
 ATTN: CHIEF ADAMS  
 616 CENTRAL AVENUE  
 ATLANTIC COUNTY  
 MINOTOLA, NJ 08341  
 USA

Your Purchase Order Number

SIM/PROBE WARRANTY  
 Any questions? Please contact:  
 MARA HERRERA

Your sales person

Don Pouliot  
 Phone: 973-398-3228

Delivery terms

FA

FREIGHT ALLOWED

Page 1 / 2

Item#	Sh. Quant.	Part no. Description	Ship via: Fed Ex Saver
0010	1 EA	MPCERT CERTIFICATION CHARGE SIMULATOR Ordered / Back ordered : 0 / 0 EA	
0020	1 EA	MPCALTP CAL CHARGE, SIM PROBE Ordered / Back ordered : 0 / 0 EA	

NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box.  
 Please call Customer Service at 866-385-5900.  
 SIM: DDBJ-0006 PROBE: DDMJP2-088

Dräger Safety Diagnostics Inc.  
 101 Technology Drive  
 Pittsburgh, PA 15275  
 Tel: (412) 788-5537  
 Fax: (412) 788-5598  
 Citizens Bank  
 Acct # 6209426615  
 Acct Name: Dräger Safety Diagnostics Inc.  
 Transil Number: 036076150  
 Federal ID Number: 84-1600159  
 Dräger Safety Diagnostics Inc.  
 P.O. Box 200337  
 Pittsburgh, PA 15251-0337  
 REMIT TO:  
 Dräger Safety Diagnostics Inc.  
 P.O. Box 200337  
 Pittsburgh, PA 15251-0337

# Packing Slip

Customer no    Order No    Order date  
150065405    10429212    09/08/2010

Packing slip no.    Ship date  
80544982    09/08/2010  
Please reference on inquiries

Page    2 / 2

Item#	sh. Qunt.	Part no. Description			
0030	1 EA	12043 HOSE PUMP-SIM W/SIM-CUVETTE HOSE 2" Ordered / Back ordered : 0 / 0 EA			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 0 / 0 EA			

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <i>Buena Borough Police Dept 616 Central Ave Minotola, NJ 08341</i>	2. Contact: <i>Sgt Petrilla</i>	4. Date: <i>8/31/10</i>
3. Phone Number: <i>609-697-2810</i>		5. County: <i>Atlantic</i>

6. Alcotest Instrument Serial Number: <i>N/A</i>	7. Simulator Component Serial Number: <i>DOB5-0006</i>	8. Temperature Probe Component Serial Number: <i>DDWJP2-088</i>
-----------------------------------------------------	-----------------------------------------------------------	--------------------------------------------------------------------

9. Reason for Service:  
*Simulator temp at 34.34 with digital temp measuring system, too high outside of tolerance.*

10. Comments:  
*Please recert simulator and temp probe together to keep expiration dates consistent. Thank you.*

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

*Tpr II Steve Dellance # 6027*  
Name & Badge Number (Print)

*Tpr II [Signature] # 6027*      *8/31/10*  
Signature      Date



STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

<b>1. Department:</b> <b>Buena Borough Police Department</b> <b>616 Central Ave.</b> <b>Minotola, NJ 08341</b>	<b>2. Contact:</b> <b>Sgt. Petrilla</b>	<b>4. Date:</b> <b>8/31/10</b>
<b>3. Phone Number:</b> <b>856-697-2810</b>		<b>5. County:</b> <b>Atlantic</b>

<b>6. Alcotest Instrument Serial Number:</b> <p style="text-align: center;">N/A</p>	<b>7. Simulator Component Serial Number:</b> <b>DDXJ S3-0209 replaced by</b> <b>DDBJ-0006</b>	<b>8. Temperature Probe Component Serial Number:</b> <b>DDWAP2-148 replaced by</b> <b>DDWJP2-088</b>
----------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------

**9. Reason for Service:**  
**Above listed components placed in service.**

**10. Comments:**

**Above listed simulator and temperature probe replaced by simulator serial # DDBJ-0006 and temperature probe serial DDWJP2-0088.**

**See Draeger packing slip.**

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

12. The above Instrument/Component is placed out of service pending further evaluation.


13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Tpr. II Steve Dellanoce #6027**  


---

Name & Badge Number (Print)


---

Signature

**8/31/10**  


---

Date

# Packing Slip

Customer no    Order No    Order date  
 150065405    10425706    08/19/2010

Packing slip no.    Ship date  
 80540960    08/19/2010  
 Please reference on Inquiries

ship to  
 150065406  
 BOROUGH OF BUENA POLICE DEPARTMENT  
 ATTN: CHIEF ADAMS  
 616 CENTRAL AVENUE  
 ATLANTIC COUNTY  
 MINOTOLA, NJ 08341  
 USA

bill to

Your Purchase Order Number  
 08/19/2010  
 SIM/PROBE WARRANTY  
 Any questions? Please contact:  
 MARA HERRERA

Customer:  
 BOROUGH OF BUENA  
 ATTN: FINANCE DEPARTMENT  
 P.O. BOX 620  
 ATLANTIC COUNTY  
 MINOTOLA, NJ 08341-0620  
 USA

Your sales person  
 Don Pouliot  
 Phone: 973-398-3228

Delivery terms  
 FA  
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Ground
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		SIM: DDXJS3-029 FAILED REPLACED FOC SER# DDBJ-0006 PROBE: DDWAP2-148 FAILED REPLACED FOC SER# DDWJP2-088
0010	1 EA	4540083 SIMULATOR, CU34 110V ELEC CTRL W/O TMETE Ordered / Back ordered : 1 / 0 EA

**Draeger Safety Diagnostics Inc.**  
 Accounting Address:  
 101 Technology Drive  
 Pittsburgh, PA 15275  
 Tel: (412) 788-5537  
 Fax: (412) 788-5598

**Remit Wire Transfers To:**  
 Citizens Bank  
 Acct. # 6209426615  
 Acct. Name: Draeger Safety Diagnostics Inc.  
 Transit Number: 036076150  
 Federal ID Number: 84-1600159

**REMIT TO:**  
 Draeger Safety Diagnostics Inc.  
 P.O. Box 200337  
 Pittsburgh, PA 15251-0337

# Packing Slip

Customer no    Order No    Order date  
150065405    10425706    08/19/2010

Packing slip no.    Ship date  
80540960    08/19/2010

Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0020	1 EA	12025 SIMULATOR, TEMPERATURE PROBE Ordered / Back ordered : 1 / 0 EA			
0030	1 EA	12043 HOSE PUMP-SIM W/SIM-CUVETTE HOSE 2" Ordered / Back ordered : 1 / 0 EA			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			

