

Ertco-Hart Digital Temperature Measuring System

REPORT OF CALIBRATION

This is to certify that the Ertco-Hart Digital Temperature Measuring System has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). Draeger Safety Diagnostics, Inc. (DSDI) recommends accuracy verification of the Ertco-Hart Digital Temperature Measuring System within 12 months of the certification date below, or sooner, according to your state specification.

DSDI equipment used for temperature verification Serial Number: 303176

Digital Units Serial Number: A18522

Probe Serial Number: 557904

Certification Date: 02/28/2006

Next Certification Due: 02/28/2007

At 34.00 °C digital unit displays 33.98 °C

Draeger Safety Diagnostics, Inc. Technician: Cent Dly