

State of New Jersey

OFFICE OF THE ATTURNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISKIN OF STATE POLICE
POST OFFICE BOX 7068
WEST TRINTON, NO 08628-0068
(609) 882-2000

JOHN J. HOFFMAN
Acting Attorney General

COLONEL JOSEPH R. PUBNIES
Superlatendens

KIN GUADAGNO

CHRIS CHRISTIP

Cinvernor

CERTIFICATION OF ANALYSIS 0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Dracger Safety, Inc.

ANALYSIS DATE: 5/06/2014

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 14D129

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1208 to 0.1217 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is April 29, 2016.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Foreusio Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alsouie, Ph.D. Research Scientist


NJSP Office of Forensic Sciences

Sworn to and subscribed before me this

day of

. 2014.

Notary





"An Internationally Accredited Agency"

Now Jerson is An Equal Opportunity Employer Princial on Recycled Proper and Recyclobic

