


STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Florence Township Police Department 711 Broad Street Florence, NJ 08518		2. Contact: Lt. Nicholas Czepiel	4. Date: 03/17/2026
		3. Phone Number: 609-754-4509	5. County: Burlington
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARMJ-0130</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>	

9. Reason for Service:
The Alcotest instrument was returned from outside evaluation and placed back in service.

10. Comments:
See Draeger Return and Repair Form.



Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<p><u>Tpr. James Fillimon #8306</u> Name & Badge Number (Print)</p>	<p><u>Tpr. James Fillimon #8306</u> Signature</p>	<p><u>3/17/2026</u> Date</p>
--	--	---

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Florence Township Police Department 711 Broad Street Florence, NJ 08518	2. Contact: Lt. Nicholas Czepiel	4. Date: 01/09/2026
	3. Phone Number: 609-499-3131	5. County: Burlington
6. Alcotest Instrument Serial Number: ARMJ-0130	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
I arrived at Florence Township Police Department to conduct the Certification of the Alcotest 9510 (Serial Number ARMJ-0130).

10. Comments:
On January 9, 2026 I attempted to conduct a Certification of the Alcotest 9510 (Serial Number ARMJ-0130) at the Florence Township Police Department, Burlington County. While performing LINERARITY(Part III) Sequential file #00038, the instrument sustained an External Standard Failure specifically during the first attempt of the .16. This revealed failures due to low electrochemical sensor (EC) results. The Operational condition of the instrument was assessed by reviewing the history of External Standard Failures.

The instrument will remain out of service at this time and returned to Draeger Inc. for evaluation.

Last Known Sequential File #: 00038

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. James Fillimon #8306 <hr style="border: 0; border-top: 1px solid black;"/> Name & Badge Number (Print)	 Signature	1/9/2026 Date
--	---	------------------

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: * TOWNSHIP OF FLORENCE

Date Received: 1/22/2026

Date given to service: 02/23/2026

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR MJ-0130

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414160	A9510 Calibration	1 EA	
4414166	Labor	.5 HR	
6808455	EC Sensor	1 EA	

Repair Notes: Replaced EC sensor due to Int-T and EC value out of tolerance. Software reinstalled.

EC SENSOR = ARUC-0499

CAL W/QC AND OPS CHECK

Service Technician JJ

Date: 02/23/2026

Delivery receipt



Customer no.
150054650

Number of report
311739468

Date of receiving Delivery receipt date
01/22/2026 02/24/2026

Please reference on inquiries

Consignee, 150054650

**TOWNSHIP OF FLORENCE
POLICE DEPARTMENT
711 BROAD ST
FLORENCE NJ 08518-2310**

Customer, 150054650

TOWNSHIP OF FLORENCE
POLICE DEPARTMENT
711 BROAD ST
FLORENCE NJ 08518-2310

Your order

Date of order:
01/12/2026
Purch.ord.no.:
WARRANTY

Branch text

Draeger contact person

ALCOTEST
800-437-2437 X5
US-Safety@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
000	Alcotest 9510 New Jersey	8327998	1028389717	ARMJ-0130
	REPAIR ALCOTEST 9510 Repair free of charge			

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road; Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS - Lockbox Operations
Attn:Draeger, Inc., Lockbox #13369
101 Woodcrest Road
Suite 201
Cherry Hill, New Jersey 08003

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no. 150054650 Number of report 311739468 Date of receiving 01/22/2026

Please reference on inquiries

Customer
TOWNSHIP OF FLORENCE
MUNICIPAL COMPLEX
711 BROAD ST
FLORENCE NJ 08518-2310

Consignee
TOWNSHIP OF FLORENCE
MUNICIPAL COMPLEX
711 BROAD ST
FLORENCE NJ 08518-2310

Your order

Date of order:
01/12/2026

Your reference:
WARRANTY

NL-Text

Your contact person

ALCOTEST
800-437-2437 X5
US-Safety@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 154999732 customer reported problem low EC sensor Eq.No. : 1028389717 SN: ARMJ-0130 Material: 8327998 Alcotest 9510 New Jersey Inv.No. : Location: customer reported problem low EC sensor results and external standard failure. EC Values are out of spec. Recommending EC Sensor replacement. New Sensor: ARUC-0499 REPAIR ALCOTEST 9510 10 YR EXTENDED WARR
	0.0 H	



Dräger

Alcotest 9510

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 9510 has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest 9510 is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864, and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your state's specifications.

Certification Date:

Serial Number:

02/23/2026

ARMJ-0130

Draeger, Inc.