

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401036

DRAEGER MEDICAL SYSTEMS INC

Sales order: 1145361076

Date: January 19, 2026

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: ± 0.002 BrAC or $\pm 2\%$ whichever is greater.

CALGAZ LOT#: 302-403530917

Manufactured Date: January 19, 2026

ETHANOL IN NITROGEN

Product Expiration: January 19, 2029

COMPONENT	PPM	(BrAC)
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	262.8	(0.101)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND49826	260.1

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.
Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY:



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC

821 Chesapeake Drive, Cambridge, MD 21613-0149

Phone: (410) 228-6400

Fax: (410) 228-4251

Delivery receipt

Customer no.:
91266631

Order no. :
2221962947

Delivery receipt no.:
177161298

Delivery receipt date:
08/26/2026

Consignee: 91267771

EASTAMPTON POLICE DEPT

1380 WOODLANE RD

US 08060 MOUNT HOLLY, NJ
USA

Your order dated:

08/06/2026

26-00850
Andrea DiTomaso

Your contact person:

ALCOTEST

Tel: 800-437-2437 X5

Fax:

Page 1 / 2

Please reference on inquiries

Payer: 91266631

EASTAMPTON POLICE DEPT

12 MANOR HOUSE CT

US 08060-437 MOUNT HOLLY, NJ
USA

Customer: 91266631

EASTAMPTON POLICE DEPT

12 MANOR HOUSE CT

US 08060-437 MOUNT HOLLY, NJ
USA

Applicable to International Orders Only: These items are controlled by the US government and authorized for export only to the country of ultimate destination for use by the ultimate consignee or end-user(s) herein identified. They may not be resold, transferred, or otherwise disposed of, to any other country or to any person other than the authorized ultimate consignee or end-user(s) either in their original form or after being incorporated into other items, without first obtaining approval from the US government or as otherwise authorized by US law and regulations.

Applicable to U.S. Orders Only: Any diversion (e.g. export/re-export, transfer, release, transshipment, etc.) contrary to U.S. law of products that are intended solely for use in the United States is prohibited.

Draeger Inc.

Our Tax ID: 23-1699096

3135 Quarry Road; Telford, PA 18969

An Equal Opportunity Employer M / F / V / H

Telephone 800-437-2437

<http://www.draeger.com>

Delivery receipt

Customer no.:
91266631

Order no.:
2221962947

Delivery receipt no.:
177161298

Delivery receipt date:
08/26/2026

Please reference on inquiries

Consignee: 91267771
Page 2 / 2

Payer: 91266631

Pos.	Quant.	Part no. Description		
900001	1 EA	4401036 Dry gas, .10% 105L (ethanol/N2)		
		Attention dangerous goods!		
		UN-Number:1956		
		BATCH:1019390013		

Number of packages

1

CARTON CONTENT SLIP

Printed: 8/27/26 16:21

SHIP FROM: Draeger Medical address 1 city state zip US				SHIPPER #: 581937639 Shipper Order #: 0177161298 Order Date: 08/06/2026 Ship Date: 08/26/2026 Warehouse: OLB2 Enterprise: DRG		
Mark For:						
SHIP TO : EASTAMPTON POLICE DEPT EASTAMPTON POLICE DEPT 1380 WOODLANE RD MOUNT HOLLY NJ 08060 US				BILL TO : EASTAMPTON POLICE DEPT EASTAMPTON POLICE DEPT 12 MANOR HOUSE CT MOUNT HOLL NJ 08060-437 US		
CARRIER : FedEx Ground						
FREIGHT TERMS : THIRDPARTY						
Special Instructions:						
Shipping Information						
Tracking Number	Customer PO#	Product No	Product Description	UOM	Qty Ord	Qty Pack
534615463496	26-00850	4401036	Dry gas, .10% 105L (ethano/N2)	EACH	1.00	1.00