# Calibrating Unit
# New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C
Location: EASTAMPTON TOWNSHIP
Serial No.: ARNK-0038

| | | | | | |
|---|---|---|---|---|---|
| Calibration File No.: | 00808 | Calib. Date: | 11/26/2024 | Calib. No.: | 00046 |
| Certification File No.: | 00809 | Cert. Date: | 11/26/2024 | Cert. No.: | 00041 |
| Linearity File No.: | 00810 | Lin. Date: | 11/26/2024 | Lin. No.: | 00041 |
| Solution File No.: | 00813 | Soln. Date: | 12/03/2024 | Soln. No.: | 00279 |
| Sequential File No.: | 00813 | File Date: | 12/03/2024 | | |

| | | | | | |
|---|---|---|---|---|---|
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDUK S3-0130 |
| Control Solution %: | 0.100% | | | Expires: | 07/23/2026 |
| Solution Control Lot: | 24260 | | | Bottle No.: | 0786 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|---|---|---|---|---|
| Ambient Air Blank | 0.000% | 02:21S | | |
| Control 1 EC | 0.099% | 02:22S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 02:22S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:23S | | |
| Control 2 EC | 0.099% | 02:23S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 02:23S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:24S | | |
| Control 3 EC | 0.099% | 02:25S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 02:25S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:25S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**
Last Name: MAERTEN First Name: RICHARD MI: J

Signature: Cpl #3407 Badge No.: 3407
Date: 12/03/2024

# Calibrating Unit
# New Standard Solution Report

| | | | | | |
|---|---|---|---|---|---|
| **Equipment** | Alcotest 7110 MKIII-C | | | Serial No.: | ARNK-0038 |
| Location: | EASTAMPTON TOWNSHIP | | | | |
| Calibration File No.: | 00767 | Calib. Date: | 01/17/2024 | Calib. No.: | 00044 |
| Certification File No.: | 00768 | Cert. Date: | 01/17/2024 | Cert. No.: | 00039 |
| Linearity File No.: | 00769 | Lin. Date: | 01/17/2024 | Lin. No.: | 00039 |
| Solution File No.: | 00790 | Soln. Date: | 07/06/2024 | Soln. No.: | 00272 |
| Sequential File No.: | 00790 | File Date: | 07/06/2024 | | |
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDUK S3-0131 |
| Control Solution %: | 0.100% | | | Expires: | 09/12/2025 |
| Solution Control Lot: | 23330 | | | Bottle No.: | 0463 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|---|---|---|---|---|
| Ambient Air Blank | 0.000% | 19:54D | | |
| Control 1 EC | 0.096% | 19:54D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.096% | 19:54D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 19:55D | | |
| Control 2 EC | 0.097% | 19:56D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 19:56D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 19:56D | | |
| Control 3 EC | 0.097% | 19:57D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 19:57D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 19:58D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MAERTEN First Name: CPL MI: R

Signature: Cpl 3407 Badge No.: 3407

Date: 07/06/2024

# Calibrating Unit
# New Standard Solution Report

| | | | | | |
|---|---|---|---|---|---|
| **Equipment** | Alcotest 7110 MKIII-C | | | Serial No.: | ARNK-0038 |
| Location: | EASTAMPTON TOWNSHIP | | | | |
| Calibration File No.: | 00767 | Calib. Date: | 01/17/2024 | Calib. No.: | 00044 |
| Certification File No.: | 00768 | Cert. Date: | 01/17/2024 | Cert. No.: | 00039 |
| Linearity File No.: | 00769 | Lin. Date: | 01/17/2024 | Lin. No.: | 00039 |
| Solution File No.: | 00784 | Soln. Date: | 06/09/2024 | Soln. No.: | 00271 |
| Sequential File No.: | 00784 | File Date: | 06/09/2024 | | |
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDUK S3-0131 |
| Control Solution %: | 0.100% | | | Expires: | 09/12/2025 |
| Solution Control Lot: | 23330 | | | Bottle No.: | 0441 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|---|---|---|---|---|
| Ambient Air Blank | 0.000% | 05:00D | | |
| Control 1 EC | 0.100% | 05:00D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 05:00D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 05:01D | | |
| Control 2 EC | 0.099% | 05:01D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 05:01D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 05:02D | | |
| Control 3 EC | 0.099% | 05:03D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 05:03D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 05:03D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:** .

Last Name: MAERTEN First Name: CPL MI: R

Signature: _____ Badge No.: 3407

Date: 06/09/2024

# Calibrating Unit
# New Standard Solution Report

| | | | | | |
|---|---|---|---|---|---|
| **Equipment** | Alcotest 7110 MKIII-C | | | Serial No.: | ARNK-0038 |
| Location: | EASTAMPTON TOWNSHIP | | | | |
| Calibration File No.: | 00767 | Calib. Date: | 01/17/2024 | Calib. No.: | 00044 |
| Certification File No.: | 00768 | Cert. Date: | 01/17/2024 | Cert. No.: | 00039 |
| Linearity File No.: | 00769 | Lin. Date: | 01/17/2024 | Lin. No.: | 00039 |
| Solution File No.: | 00780 | Soln. Date: | 05/12/2024 | Soln. No.: | 00270 |
| Sequential File No.: | 00780 | File Date: | 05/12/2024 | | |
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDUK S3-0131 |
| Control Solution %: | 0.100% | | | Expires: | 09/12/2025 |
| Solution Control Lot: | 23330 | | | Bottle No.: | 0366 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|---|---|---|---|---|
| Ambient Air Blank | 0.000% | 17:19D | | |
| Control 1 EC | 0.101% | 17:19D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 17:19D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:20D | | |
| Control 2 EC | 0.100% | 17:21D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 17:21D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:21D | | |
| Control 3 EC | 0.100% | 17:22D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 17:22D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:23D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MAERTEN First Name: CPL MI: R

Signature: GM #3467 Badge No.: 3407

Date: 05/12/2024

# Calibrating Unit
# New Standard Solution Report

| | | | | | |
|---|---|---|---|---|---|
| **Equipment** | Alcotest 7110 MKIII-C | | | Serial No.: | ARNK-0038 |
| Location: | EASTAMPTON TOWNSHIP | | | | |
| Calibration File No.: | 00767 | Calib. Date: | 01/17/2024 | Calib. No.: | 00044 |
| Certification File No.: | 00768 | Cert. Date: | 01/17/2024 | Cert. No.: | 00039 |
| Linearity File No.: | 00769 | Lin. Date: | 01/17/2024 | Lin. No.: | 00039 |
| Solution File No.: | 00779 | Soln. Date: | 04/14/2024 | Soln. No.: | 00269 |
| Sequential File No.: | 00779 | File Date: | 04/14/2024 | | |
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDUK S3-0131 |
| Control Solution %: | 0.100% | | | Expires: | 09/12/2025 |
| Solution Control Lot: | 23330 | | | Bottle No.: | 0180 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|---|---|---|---|---|
| Ambient Air Blank | 0.000% | 02:59D | | |
| Control 1 EC | 0.098% | 02:59D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 02:59D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 03:00D | | |
| Control 2 EC | 0.099% | 03:01D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 03:01D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 03:01D | | |
| Control 3 EC | 0.099% | 03:02D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 03:02D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 03:03D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MAERTEN First Name: RICHARD MI: J

Signature: CpI 2 #34707 Badge No.: 3407

Date: 04/14/2024

# Calibrating Unit
# New Standard Solution Report

| | | | | | |
|---|---|---|---|---|---|
| **Equipment** | Alcotest 7110 MKIII-C | | | Serial No.: | ARNK-0038 |
| Location: | EASTAMPTON TOWNSHIP | | | | |
| Calibration File No.: | 00767 | Calib. Date: | 01/17/2024 | Calib. No.: | 00044 |
| Certification File No.: | 00768 | Cert. Date: | 01/17/2024 | Cert. No.: | 00039 |
| Linearity File No.: | 00769 | Lin. Date: | 01/17/2024 | Lin. No.: | 00039 |
| Solution File No.: | 00778 | Soln. Date: | 03/12/2024 | Soln. No.: | 00268 |
| Sequential File No.: | 00778 | File Date: | 03/12/2024 | | |
| Calibrating Unit: | **WET** | Model No.: | CU-34 | Serial No.: | DDUK S3-0131 |
| Control Solution %: | 0.100% | | | Expires: | 09/12/2025 |
| Solution Control Lot: | 23330 | | | Bottle No.: | 0243 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|---|---|---|---|---|
| Ambient Air Blank | 0.000% | 18:15S | | |
| Control 1 EC | 0.099% | 18:16S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 18:16S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:16S | | |
| Control 2 EC | 0.099% | 18:17S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 18:17S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:18S | | |
| Control 3 EC | 0.099% | 18:18S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 18:18S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:19S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MAERTEN First Name: RICHARD MI: J

Signature: 

Badge No.: 3407

Date: 03/12/2024

# Calibrating Unit
# New Standard Solution Report

| | | | | | |
|---|---|---|---|---|---|
| **Equipment** | Alcotest 7110 MKIII-C | | | Serial No.: | ARNK-0038 |
| Location: | EASTAMPTON TOWNSHIP | | | | |
| Calibration File No.: | 00767 | Calib. Date: | 01/17/2024 | Calib. No.: | 00044 |
| Certification File No.: | 00768 | Cert. Date: | 01/17/2024 | Cert. No.: | 00039 |
| Linearity File No.: | 00769 | Lin. Date: | 01/17/2024 | Lin. No.: | 00039 |
| Solution File No.: | 00773 | Soln. Date: | 02/13/2024 | Soln. No.: | 00267 |
| Sequential File No.: | 00773 | File Date: | 02/13/2024 | | |
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDUK S3-0131 |
| Control Solution %: | 0.100% | | | Expires: | 09/12/2025 |
| Solution Control Lot: | 23330 | | | Bottle No.: | 0244 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|---|---|---|---|---|
| Ambient Air Blank | 0.000% | 01:06S | | |
| Control 1 EC | 0.099% | 01:06S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 01:06S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:07S | | |
| Control 2 EC | 0.099% | 01:08S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 01:08S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:08S | | |
| Control 3 EC | 0.099% | 01:09S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 01:09S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:10S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MAERTEN First Name: CPL MI: R

Signature: Cpl #3407 Badge No.: 3407

Date: 02/13/2024