Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARNK-0038
Location:	EASTAMPTON TOWNSHIP				
Calibration File No.:	00589	Calib. Date:	08/29/2018	Calib. No.:	00031
Certification File No.:	00590	Cert. Date:	08/29/2018	Cert. No.:	00027
Linearity File No.:	00591	Lin. Date:	08/29/2018	Lin. No.:	00027
Solution File No.:	00605	Soln. Date:	02/19/2019	Soln. No.:	00198
Sequential File No.:	00605	File Date:	02/19/2019		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUK S3-0130
Control Solution %:	0.100%			Expires:	10/17/2019
Solution Control Lot:	17360			Bottle No.:	1315

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:34S		
Control 1 EC	0.103%	23:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	23:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:35S		
Control 2 EC	0.102%	23:36S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	23:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:36S		
Control 3 EC	0.102%	23:37S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	23:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GRAHAM First Name: DANIEL MI: -

Signature: #340 Badge No.: 3406

Date: 02/19/2019

Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C — Serial No.: ARNK-0038
Location: EASTAMPTON TOWNSHIP

Calibration File No.:	00589	Calib. Date:	08/29/2018	Calib. No.:	00031
Certification File No.:	00590	Cert. Date:	08/29/2018	Cert. No.:	00027
Linearity File No.:	00591	Lin. Date:	08/29/2018	Lin. No.:	00027
Solution File No.:	00603	Soln. Date:	01/17/2019	Soln. No.:	00197
Sequential File No.:	00603	File Date:	01/17/2019		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUK S3-0130
Control Solution %:	0.100%			Expires:	10/17/2019
Solution Control Lot:	17360			Bottle No.:	0526

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:11S		
Control 1 EC	0.101%	20:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:13S		
Control 2 EC	0.102%	20:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:14S		
Control 3 EC	0.102%	20:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: GRAHAM First Name: CPL MI: D
Signature: CPL #3406 Badge No.: 3406
Date: 01/17/2019

Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARNK-0038
Location:	EASTAMPTON TOWNSHIP				
Calibration File No.:	00589	Calib. Date:	08/29/2018	Calib. No.:	00031
Certification File No.:	00590	Cert. Date:	08/29/2018	Cert. No.:	00027
Linearity File No.:	00591	Lin. Date:	08/29/2018	Lin. No.:	00027
Solution File No.:	00602	Soln. Date:	12/20/2018	Soln. No.:	00196
Sequential File No.:	00602	File Date:	12/20/2018		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUK S3-0130
Control Solution %:	0.100%			Expires:	10/17/2019
Solution Control Lot:	17360			Bottle No.:	1340

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:31S		
Control 1 EC	0.102%	09:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:33S		
Control 2 EC	0.101%	09:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:34S		
Control 3 EC	0.101%	09:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	09:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GRAHAM First Name: CPL MI: D

Signature: Cpl #3406 Badge No.: 3406

Date: 12/20/2018

Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARNK-0038
Location:	EASTAMPTON TOWNSHIP				
Calibration File No.:	00589	Calib. Date:	08/29/2018	Calib. No.:	00031
Certification File No.:	00590	Cert. Date:	08/29/2018	Cert. No.:	00027
Linearity File No.:	00591	Lin. Date:	08/29/2018	Lin. No.:	00027
Solution File No.:	00600	Soln. Date:	11/21/2018	Soln. No.:	00195
Sequential File No.:	00600	File Date:	11/21/2018		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUK S3-0130
Control Solution %:	0.100%			Expires:	10/17/2019
Solution Control Lot:	17360			Bottle No.:	1341

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:37S		
Control 1 EC	0.102%	20:38S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	20:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:39S		
Control 2 EC	0.102%	20:39S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	20:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:40S		
Control 3 EC	0.103%	20:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	20:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GRAHAM First Name: CPL MI: D

Signature: Badge No.: 3406

Date: 11/21/2018

Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C — Serial No.: ARNK-0038
Location: EASTAMPTON TOWNSHIP

Calibration File No.:	00589	Calib. Date:	08/29/2018	Calib. No.:	00031
Certification File No.:	00590	Cert. Date:	08/29/2018	Cert. No.:	00027
Linearity File No.:	00591	Lin. Date:	08/29/2018	Lin. No.:	00027
Solution File No.:	00598	Soln. Date:	10/26/2018	Soln. No.:	00194
Sequential File No.:	00598	File Date:	10/26/2018		

Calibrating Unit: WET — Model No.: CU-34 — Serial No.: DDUK S3-0130
Control Solution %: 0.100% — Expires: 10/17/2019
Solution Control Lot: 17360 — Bottle No.: 1363

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:23D		
Control 1 EC	0.101%	02:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	02:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:24D		
Control 2 EC	0.101%	02:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	02:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:25D		
Control 3 EC	0.101%	02:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: SHEPARD — First Name: MICHAEL — MI: H
Signature: 
Badge No.: 3404
Date: 10/26/2018

Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C — Serial No.: ARNK-0038
Location: EASTAMPTON TOWNSHIP

Calibration File No.:	00589	Calib. Date:	08/29/2018	Calib. No.:	00031
Certification File No.:	00590	Cert. Date:	08/29/2018	Cert. No.:	00027
Linearity File No.:	00591	Lin. Date:	08/29/2018	Lin. No.:	00027
Solution File No.:	00594	Soln. Date:	09/27/2018	Soln. No.:	00193
Sequential File No.:	00594	File Date:	09/27/2018		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUK S3-0130
Control Solution %:	0.100%			Expires:	03/20/2019
Solution Control Lot:	17110			Bottle No.:	0530

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:08D		
Control 1 EC	0.103%	20:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	20:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:09D		
Control 2 EC	0.102%	20:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	20:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:10D		
Control 3 EC	0.102%	20:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	20:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: GRAHAM First Name: PTL MI: D
Signature: PTL #3409 Badge No.: 3409
Date: 09/27/2018

Alcotest 7110 Calibration Record

Equipment Alcotest 7110 MKIII-C — Serial No.: ARNK-0038

Location: EASTAMPTON TOWNSHIP

Calibration File No.:	00589	Calib. Date:	08/29/2018	Calib. No.:	00031
Certification File No.:	00584	Cert. Date:	12/19/2017	Cert. No.:	00026
Linearity File No.:	00585	Lin. Date:	12/19/2017	Lin. No.:	00026
Solution File No.:	00588	Soln. Date:	08/02/2018	Soln. No.:	00191
Sequential File No.:	00589	File Date:	08/29/2018		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUK S3-0130
Control Solution %:	0.100%			Expires:	08/07/2019
Solution Control Lot:	17230			Bottle No.:	1384

Coordinator

Last Name: WATSON — First Name: MATTHEW — MI: R

Signature: Tpr. I Matthew R. Watson #7078 — Badge No.: 7078

Date: 08/29/2018

*Black Key Temperature Probe Serial.....# DDLBP3-0098 MRW

*Digital NIST Temperature Measuring System Serial.....# 170428362 MRW

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110,"as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0038

Location: EASTAMPTON TOWNSHIP

Calibration File No.:	00589	Calib. Date:	08/29/2018	Calib. No.:	00031
Certification File No.:	00590	Cert. Date:	08/29/2018	Cert. No.:	00027
Linearity File No.:	00585	Lin. Date:	12/19/2017	Lin. No.:	00026
Solution File No.:	00588	Soln. Date:	08/02/2018	Soln. No.:	00191
Sequential File No.:	00590	File Date:	08/29/2018		

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0130

Control Solution %: 0.100% Expires: 08/07/2019

Solution Control Lot: 17230 Bottle No.: 1384

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:22D		
Control 1 EC	0.099%	08:22D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:22D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23D		
Control 2 EC	0.099%	08:24D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:24D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:24D		
Control 3 EC	0.100%	08:25D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:25D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:25D		

All tests within acceptable tolerance.

Coordinator

Last Name: WATSON First Name: MATTHEW MI: R

Signature: Tpr. I Matthew R. Watson #7078 Badge No.: 7078

Date: 08/29/2018

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0038

Location: EASTAMPTON TOWNSHIP

Calibration File No.:	00589	Calib. Date:	08/29/2018	Calib. No.:	00031
Certification File No.:	00590	Cert. Date:	08/29/2018	Cert. No.:	00027
Linearity File No.:	00591	Lin. Date:	08/29/2018	Lin. No.:	00027
Solution File No.:	00588	Soln. Date:	08/02/2018	Soln. No.:	00191
Sequential File No.:	00591	File Date:	08/29/2018		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXD S3-0187
Control Solution %:	0.040%			Expires:	08/10/2019
Solution Control Lot:	17240			Bottle No.:	1339

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWF S3-0223
Control Solution %:	0.080%			Expires:	08/15/2019
Solution Control Lot:	17250			Bottle No.:	0391

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWF S3-0225
Control Solution %:	0.160%			Expires:	08/21/2019
Solution Control Lot:	17260			Bottle No.:	1207

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:42D		
Control 1 EC	0.041%	08:42D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.041%	08:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:44D		
Control 2 EC	0.041%	08:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.041%	08:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46D		
Control 3 EC	0.081%	08:46D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.080%	08:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48D		
Control 4 EC	0.081%	08:48D	34.0°C	*** TEST PASSED ***
Control 4 IR	0.080%	08:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:49D		
Control 5 EC	0.161%	08:50D	33.9°C	*** TEST PASSED ***
Control 5 IR	0.161%	08:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:51D		
Control 6 EC	0.162%	08:52D	33.9°C	*** TEST PASSED ***
Control 6 IR	0.161%	08:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53D		

All tests within acceptable tolerance.

Coordinator

Last Name: WATSON First Name: MATTHEW MI: R

Signature: Tpr. I Matthew M. Watson #7078 Badge No.: 7078

Date: 08/29/2018

Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C — Serial No.: ARNK-0038
Location: EASTAMPTON TOWNSHIP

Calibration File No.:	00589	Calib. Date:	08/29/2018	Calib. No.:	00031
Certification File No.:	00590	Cert. Date:	08/29/2018	Cert. No.:	00027
Linearity File No.:	00591	Lin. Date:	08/29/2018	Lin. No.:	00027
Solution File No.:	00592	Soln. Date:	08/29/2018	Soln. No.:	00192
Sequential File No.:	00592	File Date:	08/29/2018		

Calibrating Unit: WET — Model No.: CU-34 — Serial No.: DDUK S3-0130
Control Solution %: 0.100% — Expires: 10/17/2019
Solution Control Lot: 17360 — Bottle No.: 0600

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:00D		
Control 1 EC	0.100%	10:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		
Control 2 EC	0.100%	10:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		
Control 3 EC	0.101%	10:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDSFP2-0438 MRW

Changed By:
Last Name: WATSON — First Name: MATTHEW — MI: R

Signature: Tpr.I Matthew R. Watson #7078 — Badge No.: 7078
Date: 08/29/2018

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- ⊘ Model: ALCOTEST CU34
- ○ Model: MARK IIA
- ○ Other: _____

Serial Number: DDXD S3-0187

Certification Date:	Technician:	Re-Certification Due Date:
7-13-18	BS	7-13-19

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- ⊘ Model: ALCOTEST CU34
- ○ Model: MARK IIA
- ○ Other: _____

Serial Number: DDWF S3-0223

Certification Date:	Technician:	Re-Certification Due Date:
7-13-18	BS	7-13-19

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- ⊘ Model: ALCOTEST CU34
- ○ Model: MARK IIA
- ○ Other: _____

Serial Number: DDWFS3-0225

Certification Date:	Technician:	Re-Certification Due Date:
7-23-18	BS	7-23-19

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with with instrumentation that is traceable to the National Institute of Standards and Technology (NIST) The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe:	Certification Date:	Next Certification Due:
DDLBP3-0098	7-31-18	7-31-19

Probe Value: 102

Draeger, Inc. BS

Calibration Certificate No. 1750.01

Calibration complies with ISO/IEC 17025, ANSI/NCSL Z540-1, and 9001

Cert. No.: 4000-8609162

Traceable® Certificate of Calibration for Digital Thermometer

Manufactured for and distributed by: VWR International, LLC, Radnor Corporate Center, Bldg 1, Ste 200, 100 Matsonford Road, Radnor, PA 19087

Instrument Identification:

Model: 61220-601 S/N: 170428362 Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath TC-231	A79341		
Thermistor Module	A27129	12/01/17	1000401760
Temperature Probe	5267	12/06/17	B6B30059
Temperature Calibration Bath TC-191	A42238		
Thermistor Module	A27129	12/01/17	1000401760
Temperature Probe	5202	12/19/17	B6B30058-1
Temperature Calibration Bath TC-218	A73332		
Thermistor Probe	5356	1/10/18	B7104024
Readout, Digital Thermometer	B5C344	3/12/18	B7314035
Temperature Calibration Bath TC-275	B16388		
Thermistor Probe	5357	1/06/18	B7104023
Readout, Digital Thermometer	B5C344	3/12/18	B7314035

Certificate Information:

Technician: 104 Procedure: CAL-06 Cal Date: 6/08/17 Due Date: 6/08/19

Test Conditions: 23.5°C 50.0 %RH 1014 mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
°C		N.A.		0.002	0.000	Y	-0.048	0.052	0.010	>4:1
°C		N.A.		25.003	25.001	Y	24.953	25.053	0.010	>4:1
°C		N.A.		50.002	50.001	Y	49.952	50.052	0.010	>4:1
°C		N.A.		100.001	99.999	Y	99.951	100.051	0.010	>4:1

This Instrument was calibrated using Instruments Traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Min = As Left Nominal(Rounded) - Tolerance; Max = As Left Nominal(Rounded) + Tolerance; Date=MM/DD/YY

Nicol Rodriguez, Quality Manager

Aaron Judice, Technical Manager

Maintaining Accuracy:

In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometers change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
Control Company is ISO 9001:2008 Quality Certified by (DNV) Det Norske Veritas, Certificate No. CERT-01805-2006-AQ-HOU-RvA.
International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).

Traceable® is a registered trademark of Control Company © 2009 Control Company

State of New Jersey

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRISTOPHER S. PORRINO
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc. **ANALYSIS DATE:** 08/24/2017

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 17230

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1202 to 0.1216 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 07, 2019.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaouie, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 24 day of August, 2017.

Notary

PETER F MURPHY IV
My Commission Expires
August 1, 2019

State of New Jersey

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRISTOPHER S. PORRINO
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.04 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0469 to 0.0499 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc. **ANALYSIS DATE:** 08/29/2017

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 17240

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0483 to 0.0489 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 10, 2019.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaouie, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 30th day of August, 2017.

Mary E. McLaughlin
Notary

MARY ELIZABETH MCLAUGHLIN
ID # 2052190
NOTARY PUBLIC
STATE OF NEW JERSEY
My Commission Expires Dec. 24, 2018

State of New Jersey

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRISTOPHER S. PORRINO
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.08 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0939 to 0.0997 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc. **ANALYSIS DATE:** 09/07/2017

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 17250

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0963 to 0.0973 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 15, 2019.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaouie, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 11 day of September, 2017.

Notary

NOTARY PUBLIC
PETER F MURPHY IV
My Commission Expires
August 1, 2019

State of New Jersey
OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

CHRISTOPHER S. PORRINO
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.16 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1878 to 0.1994 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc. **ANALYSIS DATE:** 09/12/2017

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 17260

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1937 to 0.1957 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 21, 2019.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaouie, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 13 day of September, 2017.

Notary

PETER F MURPHY IV
My Commission Expires
August 1, 2019

State of New Jersey

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRISTOPHER S. PORRINO
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc. **ANALYSIS DATE:** 03/29/2017

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 17110

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1211 to 0.1231 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is March 20, 2019.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaouie, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 30th day of March, 2017.

Notary

MARY ELIZABETH MCLAUGHLIN
ID # 2052190
NOTARY PUBLIC
STATE OF NEW JERSEY
My Commission Expires Dec. 24, 2018

DEPARTMENT OF
Law and Public Safety
This is to certify that

Matthew R. Watson
New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C A METHOD TO DETERMINE INTOXICATION.
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 19th DAY OF August TWO THOUSAND AND Ten

SUPERINTENDENT
NEW JERSEY STATE POLICE

ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES _____

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	11-8-12	GCPA	Lum Lum
2.	7/14/15	CMPA	Adam Stanley
3.	3/23/17	LaKehurst	Michelle Smedley
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev. 03/10)

DEPARTMENT OF
Law and Public Safety
This is to certify that

Matthew R. Watson
Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C A METHOD TO DETERMINE INTOXICATION.
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 8th DAY OF June TWO THOUSAND AND Sixteen

SUPERINTENDENT
NEW JERSEY STATE POLICE

ACTING ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES _____

	DATE	Refresher Course PLACE	INSTRUCTOR
1.			
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev. 06/13)

Dräger

Alcotest® 7110 MKIII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864 and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date: 4-23-15

SERIAL NUMBER: ARNK-0038

Draeger Safety Diagnostics, Inc. BC

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- ⊘ Model: ALCOTEST CU34
- ○ Model: MARK IIA
- ○ Other: _____

Serial Number: DDUK53-0130

Certification Date: 6-29-18

Technician: BS

Re-Certification Due Date: 6-29-19

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe: DDSFP2-0438

Certification Date: 6-29-18

Next Certification Due: 6-29-19

Probe Value: 103

Draeger, Inc. BS