STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF CONSUMER AFFAIRS
OFFICE OF WEIGHTS AND MEASURES
1261 ROUTES 1 & 9 SOUTH
AVENEL, NEW JERSEY 07001

TEST NO. NJ171195 DATE: MAY 26, 2017

CERTIFICATE FOR EASTAMPTON TOWNSHIP POLICE DEPT.
725 SMITHVILLE ROAD
MT HOLLY, NEW JERSEY 08060

This certifies that Laser Speed Detection Instrument, Serial Number TJ002908, has been compared with standards of the State of New Jersey in possession of the State Superintendent of Weights and Measures. The error found in the time base component will result in an error of less than 0.030 MPH at any speed.

David Freed
Acting State Superintendent

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF CONSUMER AFFAIRS
OFFICE OF WEIGHTS AND MEASURES
1261 ROUTES 1 & 9 SOUTH
AVENEL, NEW JERSEY 07001

Date: June 5, 2015

EASTAMPTON TOWNSHIP POLICE DEPT.
725 SMITHVILLE ROAD
EASTAMPTON, NJ 08060

Subject: TEST REPORT - NJ150837 - LASER SPEED GUN DISTANCE CALIBRATION COURSE

TEST OF ACCURACY ONLY

The course(s) which you submitted to the New Jersey State Metrology Laboratory were/was compared with the State Standard (NIST Test Number 821/263710-00) on this date.

DESIGNATION	ACTUAL	ID MARKS
CENTER OF BASE POST TO:		ET PD LC
CENTER OF FIRST TARGET (1)	50 FEET (+/- <1/8")	
CENTER OF SECOND TARGET (2)	75 FEET (+/- <3/16")	

This certificate shall not be reproduced except in full without the written permission of the State Superintendent of Weights and Measures.

David Freed
Acting State Superintendent